



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-006512</i>	Instrument Location <i>GFSO - Jail</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *Installed after return from State Lab.*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <i>06-19-24</i>
-------------------------------	-------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>19 June 2024</b>
---------------------------	---------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006512  
Location = GFSO                8164.14.00 09/16  
06/19/2024                      08:55

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                28423080A3  
Standard Cyl #?                12  
Standard Expiration?            11/05/2025  
Oper No?                        100452

Flow Cal. Date:                08/20/2015  
Slope                            680  
Intercept                       -660856

IR Calibration Date:            08/21/2015  
                                  3um                            9um

-----  
0th Coef(\*100):                -23051                        -19261  
1st Coef(\*100):                266836                        137135  
2nd Coef(\*100):                3974                            1466  
H2O adj(mg/l\*10k):            785                            487

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
NATHAN MOEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006512  
Location = GFSO                              8164.14.00 09/16  
06/19/2024    08:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:56
02 Std. Gas	0.081	08:56
03 Room Air	0.000	08:57
04 Std. Gas	0.081	08:57
05 Room Air	0.000	08:58
06 Std. Gas	0.081	08:58
07 Room Air	0.000	08:59

Lot No = 28423080A3  
Cyl No = 12  
Exp Date = 11/05/2025  
County = 18                                      Oper No. = 100452

  
Operator Signature  
NATHAN MOEN

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = GFSO      8164.14.00 09/16  
06/19/2024      09:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:05
02 Std. Gas	0.081	09:05
03 Room Air	RFI*	09:05
04 Room Air	0.000	09:06

\*Invalid Test  
Inhibited - RFI

Lot No = 28423080A3  
Cyl No = 12  
Exp Date = 11/05/2025  
County = 18      Oper No. = 100452

  
Operator Signature  
NATHAN MOEN

Remarks:

Form 106-I8000