



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number	00-00947	Instrument Location	1201 14th St. SE, Ste C. Watford City, ND 58054
Reason for Install/Repair			
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____			

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).

AEN 6. Run Tests:

- A. Print Test (Level 1, Function P).
- B. ACA Test (Level 1, Function C).

AEN C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).

7. Repair and/or Maintenance Performed (if any): NIA

8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.

10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature	<i>[Signature]</i>	Date	06/24/24
---------------------------	--------------------	------	----------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	<i>[Signature]</i>	Certified Date	25 June 2024
-----------------------	--------------------	----------------	--------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MCSO 8164.14.00 09/16
06/24/2024 14:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.077	14:50
03 Room Air	0.000	14:50
04 Std. Gas	0.077	14:50
05 Room Air	0.000	14:51
06 Std. Gas	0.078	14:51
07 Room Air	0.000	14:52

Lot No = 14323080A4
Cyl No = 15
Exp Date = 06/05/2025
County = 27 Oper No. = 132007

Kari Stuart #3410

Operator Signature
KARI STUART

Remarks:

ACA test

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{ss}

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 24th day of June, 2024

Kari Stuart #3410
Mckenzie County Sheriff's Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MCSO 8164.14.00 09/16
06/24/2024 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	14:53
02 Room Air	RFI*	14:53

*Invalid Test
Inhibited - RFI

Lot No = 14323080A4
Cyl No = 15
Exp Date = 06/05/2025
County = 27 Oper No. = 132007

Kari Stuart #3416

Operator Signature
KARI STUART

Remarks: *RFI test*

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE } ss

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 24th day of JUNE, 2024
Kari Stuart #3416
McKenzie County Sheriff's Department