Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer 8000 — Alcohol Analyzer
North Dakota Model — SN 80-003061
Location = TOXL — 8164.14.00 09/16
~~06/18/2024~~ 06/20/2024 — 12:12

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Diagnostic. This was ran on 6/20/2024, not 6/18/24

Form 106-I8000