



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <b>BROST, CUTLER #3255</b>		
Location <b>DUNN COUNTY SHERIFFS OFFICE</b>	Intoxilyzer® Serial Number <b>80-005937</b>	
Gas Lot Number <b>221313</b>	Gas Cylinder Number <b>020</b>	Gas Expiration Date <b>2/5/2025</b>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- aen*  2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. **0.079** AC
  - B. **0.078** AC
  - C. **0.077** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> #3255	Date <b>6/25/24</b>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <b>7/3/24</b>

*Incorrect location code initially. Contacted Cutler Brost and asked to change code to DUNN. He sent the attached print test to show the correct location code. -AEN 7/3/24*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005937  
Location = DCSO      8164.14.00 09/16  
06/25/2024      17:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:06
02 Std. Gas	0.078	17:06
03 Room Air	0.000	17:07
04 Std. Gas	0.078	17:07
05 Room Air	0.000	17:08
06 Std. Gas	0.077	17:08
07 Room Air	0.000	17:09

Lot No = 221313  
Cyl No = 020  
Exp Date = 02/05/2025  
County = 13      Oper No. = 133300

  
\_\_\_\_\_  
Operator Signature  
CUTLER BROST

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005937  
Location = DUNN                    8164.14.00 09/16  
07/02/2024                            16:24

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-==|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-==|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            221313  
Standard Cyl #?            020  
Standard Expiration?        02/05/2025  
Oper No?                    133300

Flow Cal. Date:            11/17/2017  
Slope                        606  
Intercept                   -534563

IR Calibration Date:        11/17/2017  
                                  3um                    9um

0th Coef(*100):	-30534	-23772
1st Coef(*100):	294336	142357
2nd Coef(*100):	-906	850
H2O adj(mg/l*10k):	205	250

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CUTLER BROST

Remarks:

Form 106-I8000