



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-005937	Instrument Location DUNN COUNTY SHERIFF'S OFFICE
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> #3255	Date 6/25/2024
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date 7/3/24
---	---------------------------------

Incorrect location code initially. Contacted Cutter Brost and asked to change code to DUNN. He sent the attached print test to show the correct location code. -AEN 7/3/24

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = DCSO 8164.14.00 09/16
06/25/2024 16:49

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 221313
Standard Cyl #? 020
Standard Expiration? 02/05/2025
Oper No? 133300

Flow Cal. Date: 11/17/2017
Slope 606
Intercept -534563

IR Calibration Date: 11/17/2017
 3um 9um

0th Coef(*100): -30534 -23772
1st Coef(*100): 294336 142357
2nd Coef(*100): -906 850
H2O adj(mg/l*10k): 205 250

***** Printer Test End *****

 #3255

Operator Signature
CUTLER BROST

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = DCSO 8164.14.00 09/16
06/25/2024 16:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:53
02 Std. Gas	0.078	16:54
03 Room Air	0.000	16:54
04 Std. Gas	0.077	16:55
05 Room Air	0.000	16:55
06 Std. Gas	0.077	16:56
07 Room Air	0.000	16:56

Lot No = 221313
Cyl No = 020
Exp Date = 02/05/2025
County = 13 Oper No. = 133300

 #3255

Operator Signature
CUTLER BROST

Remarks:


Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = DCSD 8164.14.00 09/16
06/25/2024 16:58

Test	AC	Time
01 Room Air	RFI*	17:00
02 Room Air	0.000	17:01

*Invalid Test
Inhibited - RFI

 #3255

Operator Signature
CUTLER BROST

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = DUNN 8164.14.00 09/16
07/02/2024 16:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-==|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-==|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 221313
Standard Cyl #? 020
Standard Expiration? 02/05/2025
Oper No? 133300

Flow Cal. Date: 11/17/2017
Slope 606
Intercept -534563

IR Calibration Date: 11/17/2017
 3um 9um

0th Coef(*100):	-30534	-23772
1st Coef(*100):	294336	142357
2nd Coef(*100):	-906	850
H2O adj(mg/l*10k):	205	250

***** Printer Test End *****



Operator Signature
CUTLER BROST

Remarks:

Form 106-I8000