



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Justin E. Cote-Kanning</i>		
Location <i>MI FL</i>	Intoxilyzer® Serial Number <i>80-005958</i>	
Gas Lot Number <i>0152408044</i>	Gas Cylinder Number <i>037</i>	Gas Expiration Date <i>03/05/2026</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.079* AC
 B. *0.078* AC
 C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Justin E. Cote-Kanning</i>	Date <i>07/18/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>Ame Nabel</i>	Date 19Jul2024

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = MIFL 8164.14.00 09/16
07/18/2024 16:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:36
02 Std. Gas	0.079	16:36
03 Room Air	0.000	16:37
04 Std. Gas	0.078	16:37
05 Room Air	0.000	16:38
06 Std. Gas	0.078	16:38
07 Room Air	0.000	16:39


Lot No = 01524080A4

Cyl No = 37

Exp Date = 03/05/2026

County = 51

Oper No. = 100913


Operator Signature
JUSTIN COTE-KANNING

Remarks:

Form 106-I8000