



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>ADAM MALAFA</i>		
Location <i>NDSU PD- SBCP</i>	Intoxilyzer® Serial Number <i>80-004191</i>	
Gas Lot Number <i>28423080A3</i>	Gas Cylinder Number <i>002</i>	Gas Expiration Date <i>11-5-2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.079* AC
 B. *0.079* AC
 C. *0.078* AC

- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Adam Malafa</i>	Date <i>8/29/2024</i>
Reviewed By (Crime Laboratory Use Only) Anna Narehood <i>Anna Narehood</i>	Date <i>03Sep2024</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = SBCP 8164.14.00 09/16
08/29/2024 18:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:48
02 Std. Gas	0.079	18:48
03 Room Air	0.000	18:49
04 Std. Gas	0.079	18:49
05 Room Air	0.000	18:50
06 Std. Gas	0.078	18:50
07 Room Air	0.000	18:51

Lot No = 28423080A3
Cyl No = 2
Exp Date = 11/05/2025

County = 09 Oper No. = 129972



Operator Signature
ADAM MALAFA

Remarks: *Install Test*

Form 106-I8000