



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-005958</i>	Instrument Location <i>Dels Lake</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <i>Replacement</i>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- aw 12/18/24*  4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> 4534	Date <i>12/13/2024</i>
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i> Anna Narehood	Certified Date <i>18 Dec 2024</i>
---	--------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = DEVL                8164.14.00 09/16  
12/13/2024                      20:19

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            14323080A4  
Standard Cyl #?            4  
Standard Expiration?       06/05/2025  
Oper No?                    132008

Flow Cal. Date:            06/21/2013  
Slope                      716  
Intercept                  -760436

IR Calibration Date:       04/07/2015  
                                    3um                      9um

-----  
0th Coef(\*100):            -17179                      -17694  
1st Coef(\*100):            259328                      130890  
2nd Coef(\*100):            3473                        1654  
H2O adj(mg/l\*10k):        425                         422

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = DEVL      8164.14.00 09/16  
12/13/2024      20:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:28
02 Std. Gas	0.081	20:28
03 Room Air	0.000	20:29
04 Std. Gas	0.081	20:29
05 Room Air	0.000	20:30
06 Std. Gas	0.081	20:30
07 Room Air	0.000	20:31

Lot No = 14323080A4  
Cyl No = 4  
Exp Date = 06/05/2025  
County = 36      Oper No. = 132008

  
Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = DEVL      8164.14.00 09/16  
12/13/2024      20:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:25
02 Std. Gas	0.081	20:25
03 Room Air	0.000	20:26
04 Std. Gas	RFI*	20:26
05 Room Air	0.000	20:27

\*Invalid Test  
Inhibited - RFI

Lot No = 14323080A4  
Cyl No = 4  
Exp Date = 06/05/2025  
County = 36      Oper No. = 132008



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000