



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>William Earl</i>		
Location <i>DEVL</i>	Intoxilyzer® Serial Number <i>80-004201</i>	
Gas Lot Number <i>1432308044</i>	Gas Cylinder Number <i>4</i>	Gas Expiration Date <i>6/5/25</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.083* AC
 - B. *0.082* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>William Earl</i>	Date <i>1/18/25</i>
Reviewed By (Crime Laboratory Use Only) Anna Narehood <i>Anna Narehood</i>	Date 27 Jan 2025


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = DEVL 8164.14.00 09/16
01/18/2025 04:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:19
02 Std. Gas	0.083	04:20
03 Room Air	0.000	04:20
04 Std. Gas	0.082	04:20
05 Room Air	0.000	04:21
06 Std. Gas	0.082	04:21
07 Room Air	0.000	04:22

Lot No = 14323080A4
Cyl No = 4
Exp Date = 06/05/2025
County = 36 Oper No. = 132008



Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000