



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-004201</i>	Instrument Location <i>DEVL</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>1/18/2025</i>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Anna Narehood [Signature]</i>	Certified Date <i>27Jan2025</i>
---	------------------------------------

*The initial print test did not contain the standard gas information. I contacted William Earl to reprint a new print test. This test is attached at the end. AEN 01/27/25*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-004201  
Location = DEVL                   8164.14.00 09/16  
01/18/2025                           04:16

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?              
Standard Cyl #?              
Standard Expiration?        
Oper No?                   132008

Flow Cal. Date:           04/25/2018  
Slope                      580  
Intercept                 -164747

IR Calibration Date:      04/26/2018

	3um	9um
0th Coef(*100):	-26833	-28328
1st Coef(*100):	272526	132392
2nd Coef(*100):	3966	1397
H2O adj(mg/l*10k):	561	514

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature

WILLIAM EARL

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = DEVL      8164.14.00 09/16  
01/18/2025      04:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:19
02 Std. Gas	0.083	04:20
03 Room Air	0.000	04:20
04 Std. Gas	0.082	04:20
05 Room Air	0.000	04:21
06 Std. Gas	0.082	04:21
07 Room Air	0.000	04:22

Lot No = 14323080A4  
Cyl No = 4  
Exp Date = 06/05/2025  
County = 36      Oper No. = 132008



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = DEVL      8164.14.00 09/16  
01/18/2025      04:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:23
02 Std. Gas	0.083	04:23
03 Room Air	0.000	04:24
04 Reference	RFI*	04:24
05 Room Air	0.000	04:25

\*Invalid Test  
Inhibited - RFI

Lot No = 14323080A4  
Cyl No = 4  
Exp Date = 06/05/2025  
County = 36      Oper No. = 132008



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = DEVL      8164.14.00 09/16  
01/27/2025      11:07

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      14323080A4  
Standard Cyl #?      4  
Standard Expiration?      06/05/2025  
Oper No?      132008

Flow Cal. Date:      04/25/2018  
Slope      580  
Intercept      -164747

IR Calibration Date:      04/26/2018

	3um	9um
0th Coef(*100):	-26833	-28328
1st Coef(*100):	272526	132392
2nd Coef(*100):	3966	1397
H2O adj(mg/l*10k):	561	514

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000