



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kenton Kossan</i>		
Location <i>Bottineau CO Sheriff's office</i>	Intoxilyzer® Serial Number <i>90-006673</i>	
Gas Lot Number <i>231284 / 01524080A41</i> <i>KDC</i>	Gas Cylinder Number <i>028</i>	Gas Expiration Date <i>3/5/2026</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0.079 AC
B. 0.080 AC
C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Kenton Kossan</i>	Date <i>01/24/2025</i>
Reviewed By (Crime Laboratory Use Only) <i>Arma Narehood</i> <i>Arma Narehood</i>	Date <i>27 Jan 2025</i>

AEN
01/27/25
see attached.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BOTT 8164.14.00 09/16
01/24/2025 11:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.079	11:11
03 Room Air	0.000	11:12
04 Std. Gas	0.080	11:12
05 Room Air	0.000	11:13
06 Std. Gas	0.079	11:13
07 Room Air	0.000	11:13

Lot No = 01524080A4
Cyl No = 028
Exp Date = 03/05/2026
County = 05 Oper No. = 010907


Operator Signature
KENTON KOSSAN

Remarks:

Form 106-I8000