



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Ormiston, Christopher N.</i>		
Location <i>CARR</i>	Intoxilyzer® Serial Number <i>80-006504</i>	
Gas Lot Number <i>01524080A4</i>	Gas Cylinder Number <i>35</i>	Gas Expiration Date <i>3/5/2026</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.082* AC
  - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>2/3/2025</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>10Feb2025</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006504  
Location = CARR      8164.14.00 09/16  
02/03/2025      23:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:20
02 Std. Gas	0.081	23:20
03 Room Air	0.000	23:21
04 Std. Gas	0.082	23:21
05 Room Air	0.000	23:22
06 Std. Gas	0.082	23:22
07 Room Air	0.000	23:22

Lot No = 01524080A4  
Cyl No = 35  
Exp Date = 03/05/2026  
County = 16      Oper No. = 131530



Operator Signature  
CHRISTOPHER ORMISTON

Remarks:

ACA Test

Form 106-I8000