





ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) ALIEN HALL		
Location KENMARE PD	Intoxilyzer® Serial Number 80-004188	
Gas Lot Number 01524080A4	Gas Cylinder Number 26	Gas Expiration Date 03-05-26

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- AEN 2. Perform an ACA Test (Level 1, Function C).
2/19/25 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.087 AC
 - B. 0.087 AC
 - C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date 2-5-25
Reviewed By (Crime Laboratory Use Only) Anna Narehood 	Date 19Feb2025

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = KMPD 8164.14.00 09/16
02/05/2025 13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Gas	0.083	13:10
03 Room Air	0.000	13:10
04 Std. Gas	0.083	13:11
05 Room Air	0.000	13:11
06 Std. Gas	0.083	13:11
07 Room Air	0.000	13:12

Lot No = 01524080A4
Cyl No = 26
Exp Date = 03/05/2026
County = 51 Oper No. = 133223



Operator Signature
ALLEN HALL

Remarks:

Form 106-I8000