



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Gabriel U/2012</i>		
Location <i>Harvey P.D.</i>	Intoxilyzer® Serial Number <i>80-006540</i>	
Gas Lot Number <i>02S24080A4</i>	Gas Cylinder Number <i>2S</i>	Gas Expiration Date <i>3/05/2026</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.083* AC
 B. *0.082* AC
 C. *0.083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Gabriel U/2012</i>	Date <i>02/22/25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i> Anna Narehood	Date 12Feb2025


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.14.00 09/16
02/12/2025 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:04
02 Std. Gas	0.083	10:04
03 Room Air	0.000	10:05
04 Std. Gas	0.082	10:05
05 Room Air	0.000	10:05
06 Std. Gas	0.083	10:06
07 Room Air	0.000	10:06

Lot No = 01524080A4
Cyl No = 25
Exp Date = 03/05/2026
County = 52 Oper No. = 132537



Operator Signature
GABRIELA URZOLA

Remarks:

Form 106-I8000

Gabriela Urzola
02/22/25