



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6490

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:

Gas Cylinder Yes or No (If Yes, Lot # 01524080A4 Cyl. # 20)

Keys Yes or No

Power Cord Yes or No

2. ☒ Download Data

3. ☒ Upload Operator File

4. ☒ Current Location Code: ASHL

5. ☒ Battery Check

Was the external battery pack replaced? Yes or No

6. ☒ O-Rings

Replaced Simulator O-Ring Yes or No

Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode

2. ☒ Breath tube heated

3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.

Time Zone: CST or MDT (Time on test records will be in time zone circled).

4. ☒ Print test (Level 1,P). Sign and attach test record.

5. ☒ Tank monitor (Level 3,D,G).

Display: 972 psi Regulator: 1000 psi

Display and Regulator $\pm$ 50 psi of each other Yes or No

Gas tank tare necessary? Yes or No

If Yes, display readings after tare (Level 3,M,C,G):

Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - Low AC (Level 1,C)

Known Value $\leq$ 0.03 AC: 0.020 AC

Sim. Ser #: MP5321

Lot #: 202311D

Exp. Date: 11/28/25

☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value $\geq$ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results $\pm$ 5% AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A4
Cylinder #: 018
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: ICS 8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
- | | | |
|----------------------|---------------------|---------------------|
| Test 1 <u>0.080</u> | Test 4 <u>0.080</u> | Test 7 <u>0.079</u> |
| Test 2 <u>0.080</u> | Test 5 <u>0.079</u> | Test 8 <u>0.080</u> |
| Test 3 <u>0.079</u> | Test 6 <u>0.080</u> | Test 9 <u>0.079</u> |
| Average <u>0.079</u> | | |
- ☒ Results $\pm$ 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

02 April 2025
Date

[Signature]
Reviewer

03 Apr 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 20
Standard Expiration? 03/05/2026
Oper No? 133237

Flow Cal. Date: 09/04/2020
Slope 649
Intercept -598817

IR Calibration Date: 09/04/2020
 3um 9um

0th Coef(*100):	-25190	-16392
1st Coef(*100):	271645	136906
2nd Coef(*100):	2533	1222
H2O adj(mg/l*10k):	460	309

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Sol.	0.021	13:34
03 Room Air	0.000	13:35
04 Std. Sol.	0.021	13:35
05 Room Air	0.000	13:36
06 Std. Sol.	0.021	13:37
07 Room Air	0.000	13:37

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check- 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Sol.	0.298	13:40
03 Room Air	0.000	13:41
04 Std. Sol.	0.303	13:41
05 Room Air	0.000	13:42
06 Std. Sol.	0.303	13:43
07 Room Air	0.000	13:43

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac check -0.300Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:45

Test	AC	Time
01 Room Air	0.000	13:46
02 *Subject Test	INT*	13:46
03 Room Air	0.000	13:47

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:47

Test	AC	Time
01 Diagnostic	OK	13:48
02 Room Air	0.000	13:49
03 *Subject Test	RFI*	13:49
04 Room Air	0.000	13:49

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.080	13:51
03 Room Air	0.000	13:52
04 Std. Gas	0.080	13:52
05 Room Air	0.000	13:53
06 Std. Gas	0.079	13:53
07 Room Air	0.000	13:54

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.080	13:55
03 Room Air	0.000	13:56
04 Std. Gas	0.079	13:56
05 Room Air	0.000	13:57
06 Std. Gas	0.080	13:57
07 Room Air	0.000	13:58

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/02/2025 13:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:59
02 Std. Gas	0.079	13:59
03 Room Air	0.000	14:00
04 Std. Gas	0.080	14:00
05 Room Air	0.000	14:01
06 Std. Gas	0.079	14:01
07 Room Air	0.000	14:02

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000