



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6506 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: TOXL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 686 psi Regulator: 700 psi
 - Display and Regulator ± 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value ≤ 0.03 AC: 0.020 AC
 - Sim. Ser #: MP5321
 - Lot #: 202311D
 - Exp. Date: 11/28/25
 - Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.300 AC
 Sim. Ser #: MP6035
 Lot #: 202402C
 Exp. Date: 2/14/26
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 1B
 Exp. Date: 6/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7352
 Lot #: ICS 8
 Exp. Date: None
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]
 Inspector Signature

02 April 2025
 Date

[Signature]
 Reviewer

03 Apr 2025
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.300
Standard Solution #? 202402C
Sim Serial #? MP6035
Oper No? 133237

Flow Cal. Date: 08/20/2015
Slope 657
Intercept -627881

IR Calibration Date: 08/20/2015
 3um 9um

0th Coef(*100): -22470 -18919
1st Coef(*100): 266505 136109
2nd Coef(*100): 3702 1644
H2O adj(mg/l*10k): 709 510

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Pnnt Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:18
02 Std. Sol.	0.021	12:18
03 Room Air	0.000	12:19
04 Std. Sol.	0.021	12:20
05 Room Air	0.000	12:20
06 Std. Sol.	0.021	12:21
07 Room Air	0.000	12:21

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Low AC Check - 0.020AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:23
02 Std. Sol.	0.300	12:24
03 Room Air	0.000	12:25
04 Std. Sol.	0.303	12:26
05 Room Air	0.000	12:26
06 Std. Sol.	0.302	12:27
07 Room Air	0.000	12:28
08 Sim Temp = 34.0°C		

Simul Ser No = MP6035
Std Sol No = 202402C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:32

Test	AC	Time
01 Room Air	0.000	12:33
02 *Subject Test	INT*	12:34
03 Room Air	0.000	12:34

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:35

Test	AC	Time
01 Diagnostic	OK	12:36
02 Room Air	0.000	12:37
03 *Subject Test	RFI*	12:37
04 Room Air	0.000	12:37

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Gas	0.081	12:39
03 Room Air	0.000	12:39
04 Std. Gas	0.081	12:40
05 Room Air	0.000	12:40
06 Std. Gas	0.080	12:40
07 Room Air	0.000	12:41

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:42
02 Std. Gas	0.081	12:42
03 Room Air	0.000	12:43
04 Std. Gas	0.081	12:43
05 Room Air	0.000	12:44
06 Std. Gas	0.081	12:44
07 Room Air	0.000	12:44

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-7
4-6 AEN
02 APR 2025

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = TOXL 8164.14.00 09/16
04/02/2025 12:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.081	12:46
03 Room Air	0.000	12:46
04 Std. Gas	0.081	12:47
05 Room Air	0.000	12:47
06 Std. Gas	0.080	12:48
07 Room Air	0.000	12:48

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #7-9

Form 106-I8000