



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3059 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # 01524080A4 Cyl. # 33)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: NS
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 998 psi Regulator: 1000 psi
Display and Regulator $\pm$ 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value $\leq$ 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 2023110
Exp. Date: 11/28/25
☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 28423080A3
Cylinder #: 37
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: ICS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.082 Test 4 0.082 Test 7 0.082
Test 2 0.081 Test 5 0.081 Test 8 0.081
Test 3 0.081 Test 6 0.081 Test 9 0.083
Average 0.081
☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

02 April 2025
Date

[Signature]
Reviewer

03 Apr 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:17

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 33
Standard Expiration? 03/05/2026
Oper No? 133237

Flow Cal. Date: 11/28/2018
Slope 589
Intercept -276335

IR Calibration Date: 12/04/2018
 3um 9um

0th Coef(*100):	-25330	-15825
1st Coef(*100):	278286	137135
2nd Coef(*100):	1992	1257
H2O adj(mg/l*10k):	525	321

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:24
02 Std. Sol.	0.020	12:26
03 Room Air	0.000	12:26
04 Std. Sol.	0.019	12:27
05 Room Air	0.000	12:27
06 Std. Sol.	0.020	12:28
07 Room Air	0.000	12:29

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:31
02 Std. Sol.	0.300	12:32
03 Room Air	0.000	12:32
04 Std. Sol.	0.301	12:33
05 Room Air	0.000	12:34
06 Std. Sol.	0.301	12:34
07 Room Air	0.000	12:35

08 Sim Temp = 34.0°C

Simul Ser No = MP6035
Std Sol No = 202402C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:36

Test	AC	Time
01 Room Air	0.000	12:37
02 *Subject Test	INT*	12:37
03 Room Air	0.000	12:38

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 37
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:39

Test	AC	Time
01 Diagnostic	OK	12:40
02 Room Air	0.000	12:41
03 *Subject Test	RFI*	12:41
04 Room Air	0.000	12:42

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 37
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.


Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Gas	0.082	12:43
03 Room Air	0.000	12:43
04 Std. Gas	0.081	12:44
05 Room Air	0.000	12:44
06 Std. Gas	0.081	12:45
07 Room Air	0.000	12:45

Lot No = 28423080A3
Cyl No = 37
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.082	12:46
03 Room Air	0.000	12:47
04 Std. Gas	0.081	12:47
05 Room Air	0.000	12:48
06 Std. Gas	0.081	12:48
07 Room Air	0.000	12:49

Lot No = 28423080A3
Cyl No = 37
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check # 4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/02/2025 12:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:54
02 Std. Gas	0.082	12:55
03 Room Air	0.000	12:55
04 Std. Gas	0.081	12:56
05 Room Air	0.000	12:56
06 Std. Gas	0.083	12:56
07 Room Air	0.000	12:57

Lot No = 28423080A3
Cyl No = 37
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #7-9

Form 106-I8000