



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6681

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:

Gas Cylinder Yes or No (If Yes, Lot # 14323080AH Cyl. # 65)

Keys Yes or No

Power Cord Yes or No

2. ☒ Download Data

3. ☒ Upload Operator File

4. ☒ Current Location Code: CARS

5. ☒ Battery Check

Was the external battery pack replaced? Yes or No

6. ☒ O-Rings

Replaced Simulator O-Ring Yes or No

Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode

2. ☒ Breath tube heated

3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.

Time Zone: CST or MDT (Time on test records will be in time zone circled).

4. ☒ Print test (Level 1,P). Sign and attach test record.

5. ☒ Tank monitor (Level 3,D,G).

Display: 974 psi Regulator: 950 psi

Display and Regulator $\pm$ 50 psi of each other Yes or No

Gas tank tare necessary? Yes or No

If Yes, display readings after tare (Level 3,M,C,G):

Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - Low AC (Level 1,C)

Known Value $\leq$ 0.03 AC: 0.020 AC

Sim. Ser #: MP5321

Lot #: 202311D

Exp. Date: 11/28/25

☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 02/14/26
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 28423080A3
Cylinder #: 41
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: 1CS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
- | | | | | | |
|--------|--------------|---------|--------------|--------|--------------|
| Test 1 | <u>0.079</u> | Test 4 | <u>0.079</u> | Test 7 | <u>0.079</u> |
| Test 2 | <u>0.079</u> | Test 5 | <u>0.080</u> | Test 8 | <u>0.079</u> |
| Test 3 | <u>0.079</u> | Test 6 | <u>0.080</u> | Test 9 | <u>0.080</u> |
| | | Average | <u>0.079</u> | | |
- ☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

03 April 2025
Date

[Signature]
Reviewer

03 Apr 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:07

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A1
Standard Cyl #? 65
Standard Expiration? 06/05/2025
Oper No? 133237

Flow Cal. Date: 05/27/2020
Slope 651
Intercept -622365

IR Calibration Date: 05/27/2020
3um 9um

0th Coef(*100):	-10420	-19573
1st Coef(*100):	269401	137403
2nd Coef(*100):	1095	923
H2O adj(mg/l*10k):	127	287

***** Printer Test End *****


Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Sol.	0.021	12:09
03 Room Air	0.000	12:09
04 Std. Sol.	0.022	12:10
05 Room Air	0.000	12:10
06 Std. Sol.	0.022	12:11
07 Room Air	0.000	12:12

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check -0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:13
02 Std. Sol.	0.299	12:14
03 Room Air	0.000	12:14
04 Std. Sol.	0.301	12:15
05 Room Air	0.000	12:15
06 Std. Sol.	0.301	12:16
07 Room Air	0.000	12:17


08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:18

Test	AC	Time
01 Room Air	0.000	12:18
02 *Subject Test	INT*	12:19
03 Room Air	0.000	12:20

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 41
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:20

Test	AC	Time
01 Diagnostic	OK	12:21
02 Room Air	RFI*	12:21
03 Room Air	0.000	12:21

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 41
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:22
02 Std. Gas	0.079	12:22
03 Room Air	0.000	12:23
04 Std. Gas	0.079	12:23
05 Room Air	0.000	12:24
06 Std. Gas	0.079	12:24
07 Room Air	0.000	12:25

Lot No = 28423080A3
Cyl No = 41
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:26
02 Std. Gas	0.079	12:27
03 Room Air	0.000	12:27
04 Std. Gas	0.080	12:28
05 Room Air	0.000	12:28
06 Std. Gas	0.080	12:29
07 Room Air	0.000	12:29

Lot No = 28423080A3
Cyl No = 41
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check # 4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = TOXL 8164.14.00 09/16
04/03/2025 12:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:30
02 Std. Gas	0.079	12:30
03 Room Air	0.000	12:31
04 Std. Gas	0.079	12:31
05 Room Air	0.000	12:32
06 Std. Gas	0.080	12:32
07 Room Air	0.000	12:33

Lot No = 28423080A3
Cyl No = 41
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000