



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4204

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: TOXL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 578 psi Regulator: 600 psi
Display and Regulator $\pm$ 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value $\leq$ 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 02/14/26
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A4
Cylinder #: 18
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: IC58
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.080 Test 4 0.081 Test 7 0.082
Test 2 0.081 Test 5 0.082 Test 8 0.082
Test 3 0.081 Test 6 0.081 Test 9 0.082
Average 0.082 0.081 den 03Apr25
☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

03 April 2025
Date

[Signature]
Reviewer

03 Apr 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 12:57

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 11
Standard Expiration? 11/05/2025
Oper No? 133237

Flow Cal. Date: 04/15/2020
Slope 657
Intercept -673032

IR Calibration Date: 06/22/2023
 3um 9um

0th Coef(*100):	-20901	-22715
1st Coef(*100):	264345	135832
2nd Coef(*100):	2957	1309
H2O adj(mg/l*10k):	448	454

***** Printer Test End *****


Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 12:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Sol.	0.020	13:02
03 Room Air	0.000	13:03
04 Std. Sol.	0.021	13:04
05 Room Air	0.000	13:04
06 Std. Sol.	0.021	13:05
07 Room Air	0.000	13:05

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:07
02 Std. Sol.	0.301	13:08
03 Room Air	0.000	13:08
04 Std. Sol.	0.305	13:09
05 Room Air	0.000	13:10
06 Std. Sol.	0.306	13:10
07 Room Air	0.000	13:11

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:12

Test	AC	Time
01 Room Air	0.000	13:13
02 *Subject Test	INT*	13:14
03 Room Air	0.000	13:14

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:15

Test	AC	Time
01 Diagnostic	OK	13:16
02 Room Air	0.000	13:16
03 Reference	RFI*	13:16
04 Room Air	0.000	13:17

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Gas	0.080	13:18
03 Room Air	0.000	13:18
04 Std. Gas	0.081	13:19
05 Room Air	0.000	13:19
06 Std. Gas	0.081	13:20
07 Room Air	0.000	13:20

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:21
02 Std. Gas	0.081	13:21
03 Room Air	0.000	13:22
04 Std. Gas	0.082	13:22
05 Room Air	0.000	13:23
06 Std. Gas	0.081	13:23
07 Room Air	0.000	13:24

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.14.00 09/16
04/03/2025 13:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Gas	0.082	13:25
03 Room Air	0.000	13:25
04 Std. Gas	0.082	13:26
05 Room Air	0.000	13:26
06 Std. Gas	0.082	13:27
07 Room Air	0.000	13:27

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #7-9

Form 106-I8000