



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Terry Dillon</i>		
Location <i>CARS</i>	Intoxilyzer® Serial Number <i>80-006681</i>	
Gas Lot Number <i>25024080A3</i>	Gas Cylinder Number <i>049</i>	Gas Expiration Date <i>10-05-2026</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.080* AC
 B. *0.079* AC
 C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>4-3-25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>15 April 2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = CARS 8164.14.00 09/16
04/03/2025 16:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:10
02 Std. Gas	0.080	16:10
03 Room Air	0.000	16:11
04 Std. Gas	0.079	16:11
05 Room Air	0.000	16:12
06 Std. Gas	0.080	16:12
07 Room Air	0.000	16:12

Lot No = 25024080
Cyl No = 049
Exp Date = 10/05/2026
County = 19 Oper No. = 132318



Operator Signature
TERRY DILLON

Remarks:

Form 106-I8000