



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Anthony Haut</i>		
Location <i>HACTC - Rugby - RUGB</i>		Intoxilyzer® Serial Number <i>80-004953</i>
Gas Lot Number <i>25024080A3</i>	Gas Cylinder Number <i>59</i>	Gas Expiration Date <i>10-05-2026</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.083* AC
 - B. *0.083* AC
 - C. *0.083* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Anthony Haut</i>	Date <i>4/3/25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>15 April 2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = RUGB 8164.14.00 09/16
04/03/2025 19:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:48
02 Std. Gas	0.083	19:48
03 Room Air	0.000	19:49
04 Std. Gas	0.083	19:49
05 Room Air	0.000	19:50
06 Std. Gas	0.083	19:50
07 Room Air	0.000	19:51

Lot No = 25024080A3
Cyl No = 58
Exp Date = 10/05/2026
County = 35

Oper No. = 132732

 4525

Operator Signature
ANTHONY HAUF

Remarks:

Form 106-I8000