




**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
SFN 59281 (06/2018)

Serial Number <b>80-006506</b>	Instrument Location <b>MANDAN</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- AEN
- ☒ 1. Surge Protector Installed/Property Grounded.
  - ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
  - ☒ 3. Breath Tube Heated.
  - ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
  - ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
  - ☒ 6. Run Tests:
    - ☒ A. Print Test (Level 1, Function P).
    - ☒ B. ACA Test (Level 1, Function C).
    - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
  - ☐ 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
  - ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
  - ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
  - ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>4/3/25</b>
--	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Anna Narehood</b> 	Certified Date <b>15April2025</b>
--	--------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006506  
Location = MAND      8164.14.00 09/16  
04/03/2025      14:26

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      14323080A1  
Standard Cyl #?      73  
Standard Expiration?      06/05/2025  
Oper No?      020611

Flow Cal. Date:      08/20/2015  
Slope      657  
Intercept      -627881

IR Calibration Date:      08/20/2015  
                                    3um                      9um

0th Coef(*100):	-22470	-18919
1st Coef(*100):	266505	136109
2nd Coef(*100):	3702	1644
H2O adj(mg/l*10k):	709	510

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006506  
Location = MAND      8164.14.00 09/16  
04/03/2025      14:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:27
02 Std. Gas	0.081	14:27
03 Room Air	0.000	14:28
04 Std. Gas	0.080	14:28
05 Room Air	0.000	14:29
06 Std. Gas	0.081	14:29
07 Room Air	0.000	14:30

Lot No = 14323080A1  
Cyl No = 73  
Exp Date = 06/05/2025  
County = 30      Oper No. = 020611

  
\_\_\_\_\_  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006506  
Location = MAND      8164.14.00 09/16  
04/03/2025      14:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	14:31
02 Room Air	RFI*	14:32

\*Invalid Test  
Inhibited - RFI

Lot No = 14323080A1  
Cyl No = 73  
Exp Date = 06/05/2025  
County = 30      Oper No. = 020611

  
Operator Signature  
TOM SHARP

Remarks:

Form 106-I8000