



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <i>80-006691</i>	Instrument Location <i>NEWT</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> 132000	Date <i>4/4/25</i>
--	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i> Anna Narehood	Certified Date <i>4/14/2025</i>
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = NEWT 8164.14.00 09/16
04/04/2025 16:21

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 9
Standard Expiration? 10/05/2026
Oper No? 132000

Flow Cal. Date: 08/28/2018
Slope 699
Intercept -660610

IR Calibration Date: 04/03/2024

	3um	9um
0th Coef(*100):	-12327	-23187
1st Coef(*100):	262018	135444
2nd Coef(*100):	2687	1323
H2O adj(mg/l*10k):	263	405

***** Printer Test End *****



Operator Signature
MARK NEW HOLY

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = NEWT 8164.14.00 09/16
04/04/2025 16:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:14
02 Std. Gas	0.078	16:15
03 Room Air	0.000	16:15
04 Std. Gas	0.078	16:16
05 Room Air	0.000	16:16
06 Std. Gas	0.077	16:16
07 Room Air	0.000	16:17

Lot No = 25024080A3
Cyl No = 9
Exp Date = 10/05/2026
County = 60 Oper No. = 132000



Operator Signature
MARK NEW HOLY

Remarks: ACA Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = NEWT 8164.14.00 09/16
04/04/2025 16:23

Test	AC	Time
01 Room Air	RFI*	16:24
02 Room Air	0.000	16:24

*Invalid Test
Inhibited - RFI

Sub Name = TEST, THOMAS T
Sub DOB = 12/31/1990
Sub Sex = Male Weight = 300
Test = DUI Cit = TEST1234
Dr. Lic. = ND/TES901234
Lot No = 25024080A3
Cyl No = 9
Expiration Date = 10/05/2026
County = 60 Oper No. = 132000



Operator Signature
MARK NEW HOLY

Remarks: *RFI Test*

Form 106-I8000