



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5945 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 14323080A4 Cyl. # 8)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: BSPD
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 905 psi Regulator: 900 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - High AC (Level 1,C)  
Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP6035  
Lot #: 202402C  
Exp. Date: 02/14/20  
☒ Results  $\pm 5\%$  AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
Known Value : 0.080 AC  
Gas Cylinder Lot #: 28423080A3  
Cylinder #: 41  
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)  
Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7352  
Lot #: ICS 8  
Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
- |        |              |         |              |        |              |
|--------|--------------|---------|--------------|--------|--------------|
| Test 1 | <u>0.081</u> | Test 4  | <u>0.080</u> | Test 7 | <u>0.080</u> |
| Test 2 | <u>0.080</u> | Test 5  | <u>0.081</u> | Test 8 | <u>0.079</u> |
| Test 3 | <u>0.081</u> | Test 6  | <u>0.080</u> | Test 9 | <u>0.080</u> |
|        |              | Average | <u>0.080</u> |        |              |
- ☒ Results  $\pm 0.005$  AC of known AC

D. Remarks/Maintenance: N/A

---

---

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

---

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

Chris Muel  
Inspector Signature

15 April 2025  
Date

Janelli Pirochona  
Reviewer

16 Apr 2025  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:09
02 Std. Sol.	0.019	12:10
03 Room Air	0.000	12:10
04 Std. Sol.	0.020	12:11
05 Room Air	0.000	12:12
06 Std. Sol.	0.020	12:12
07 Room Air	0.000	12:13

08 Sim Temp = 34.0°C

Simul Ser No = MP5321  
Std Sol No = 202311D  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:19

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:19
02 Std. Sol.	0.305	12:20
03 Room Air	0.000	12:21
04 Std. Sol.	0.305	12:22
05 Room Air	0.000	12:22
06 Std. Sol.	0.307	12:23
07 Room Air	0.000	12:23

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:25

Test	AC	Time
01 Room Air	0.000	12:26
02 *Subject Test	INT*	12:27
03 Room Air	0.000	12:28

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = DUI      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 41  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:31

Test	AC	Time
01 Diagnostic	OK	12:31
02 Room Air	0.000	12:32
03 *Subject Test	RFI*	12:32
04 Room Air	0.000	12:33

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 41  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:34
02 Std. Gas	0.081	12:34
03 Room Air	0.000	12:35
04 Std. Gas	0.080	12:35
05 Room Air	0.000	12:35
06 Std. Gas	0.081	12:36
07 Room Air	0.000	12:36

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Gas	0.080	12:38
03 Room Air	0.000	12:39
04 Std. Gas	0.081	12:39
05 Room Air	0.000	12:39
06 Std. Gas	0.080	12:40
07 Room Air	0.000	12:40

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks:

Dry cal check # 4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.14.00 09/16  
04/15/2025      12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Gas	0.080	12:42
03 Room Air	0.000	12:42
04 Std. Gas	0.079	12:43
05 Room Air	0.000	12:43
06 Std. Gas	0.080	12:44
07 Room Air	0.000	12:44

Lot No = 28423080A3  
Cyl No = 41  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000