

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number	80-006504	Instrument Location	CARR
Reason for Install/Repair			
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change			
<input type="checkbox"/> Other (Specify) _____			


Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☐ 7. Repair and/or Maintenance Performed (if any): N/A
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature	 1621	Date	4/16/2025
---------------------------	------------------------------------------------------------------------------------------	------	-----------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Anna Narehood		Certified Date	24April2025
-----------------------	---------------	--------------------------------------------------------------------------------------	----------------	-------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = CARR 8164.14.00 09/16
04/16/2025 18:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 35
Standard Expiration? 03/05/2026
Oper No? 131530

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 05/31/2017
3um 9um

0th Coef(*100):	-21488	-16160
1st Coef(*100):	265581	137324
2nd Coef(*100):	2507	1149
H2O adj(mg/l*10k):	519	348

***** Printer Test End *****



Operator Signature
CHRISTOPHER ORMISTON

Remarks:

Print Test

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = CARR 8164.14.00 09/16
04/16/2025 18:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:35
02 Std. Gas	0.081	18:35
03 Room Air	0.000	18:36
04 Std. Gas	0.081	18:36
05 Room Air	0.000	18:37
06 Std. Gas	0.081	18:37
07 Room Air	0.000	18:38

Lot No = 01524080A4
Cyl No = 35
Exp Date = 03/05/2026
County = 16 Oper No. = 131530



Operator Signature
CHRISTOPHER ORMISTON

Remarks:

ACA Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = CARR 8164.14.00 09/16
04/16/2025 18:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	18:43
02 Room Air	0.000	18:43

*Invalid Test
Inhibited - RFI

Lot No = 01524080A4
Cyl No = 35
Exp Date = 03/05/2026
County = 16 Oper No. = 131530



Operator Signature
CHRISTOPHER ORMISTON

Remarks:

RFI Test

Form 106-18000