



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80-003071	Instrument Location Bowman County Sheriff's Office
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Install after Annual Inspection at Crime Lab	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests: **AEN**
4/24/2025
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): **NA**
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature Rep David Remundin 4/6/25	Date 20 April 2025
---	------------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Narehood	Certified Date 24 April 2025
---	--

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = BOWM 8164.14.00 09/16
04/20/2025 10:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 7
Standard Expiration? 11/05/2025
Oper No? 134062

Flow Cal. Date: 07/09/2010
Slope 806
Intercept -994737

IR Calibration Date: 04/10/2015
 3um 9um

0th Coef(*100):	-11328	-10938
1st Coef(*100):	268087	134278
2nd Coef(*100):	3093	1368
H2O adj(mg/l*10k):	242	233

***** Printer Test End *****

Ref Garret Roemmich 4163

Operator Signature
GARRET ROEMMICH

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = BOWM 8164.14.00 09/16
04/20/2025 10:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Gas	0.079	10:45
03 Room Air	0.000	10:45
04 Std. Gas	0.079	10:46
05 Room Air	0.000	10:46
06 Std. Gas	0.079	10:47
07 Room Air	0.000	10:47

Lot No = 28423080A3
Cyl No = 7
Exp Date = 11/05/2025
County = 06 Oper No. = 134062

 4/63

Operator Signature
GARRET ROEMMICH

Remarks: *ACA Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = BOWM 8164.14.00 09/16
04/20/2025 10:51

Test	AC	Time
01 Room Air	RFI*	10:55
02 Room Air	0.000	10:55

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA
Sub DOB = 04/20/2000
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 28423080A3
Cyl No = 7
Expiration Date = 11/05/2025
County = 06 Oper No. = 134062

Deputy Sheriff Roemmich 4163
Operator Signature
GARRET ROEMMICH

Remarks: RFI Test

Form 106-I8000