



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
SFN 59281 (06/2018)

Serial Number <u>80-004948</u>	Instrument Location <u>Watford City PD</u>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
  - ☒ A. Print Test (Level 1, Function P).
  - ☒ B. ACA Test (Level 1, Function C).
  - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>Melissa Carl</u>	Date <u>4-20-25</u>
---	---------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>Anna Narehood</u>	Certified Date <u>21 April 2025</u>
---	--

STATE OF NORTH DAKOTA  
COUNTY OF MCKENZIE

} ss

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 20 day of April 2025  
Melissa Carl  
Watford City Police Department

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = WCPD      8164.14.00 09/16  
04/20/2025      06:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 28423080A3  
Standard Cyl #? 1  
Standard Expiration? 11/05/2025  
Oper No? 133050  
  
Flow Cal. Date: 08/28/2018  
Slope 642  
Intercept -606282

IR Calibration Date: 04/01/2024  
3um 9um

0th Coef(*100):	-23587	-22037
1st Coef(*100):	256489	131821
2nd Coef(*100):	3363	1546
H2O adj(mg/l*10k):	339	298

\*\*\*\*\* Printer Test End \*\*\*\*\*

*Matthew Earl*

Operator Signature  
MATTHEW EARL

Remarks:

*Install*

Form 106-I8000

STATE OF NORTH DAKOTA  
COUNTY OF MCKENZIE

} ss

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 20 day of April 2025

*Matthew Earl*  
Watford City Police Department

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = WCPD      8164.14.00 09/16  
04/20/2025      13:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Gas	0.078	13:22
03 Room Air	0.000	13:23
04 Std. Gas	0.077	13:23
05 Room Air	0.000	13:24
06 Std. Gas	0.079	13:24
07 Room Air	0.000	13:25

Lot No = 28423080A3

Cyl No = 1

Exp Date = 11/05/2025

County = 27

Oper No. = 133050

  
Operator Signature  
MATTHEW EARL


Remarks: *Inst 11*

Form 106-I8000

STATE OF NORTH DAKOTA } ss  
COUNTY OF MCKENZIE

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 20 day of April 2025

  
Watford City Police Department

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = WCPD      8164.14.00 09/16  
04/20/2025      13:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	13:27
02 Room Air	0.000	13:27

\*Invalid Test  
Inhibited - RFI

Lot No = 28423080A3  
Cyl No = 1  
Exp Date = 11/05/2025  
County = 27      Oper No. = 133050



Operator Signature  
MATTHEW EARL


Remarks:

*INST 91*

Form 106-I8000

STATE OF NORTH DAKOTA } ss  
COUNTY OF MCKENZIE

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 20 day of April 2025  
  
Watford City Police Department