



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number <i>80-004936</i>	Instrument Location <i>8m DC</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): *N/A*
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>04/21/2025</i>
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Anna Narehood</i>	Certified Date <i>25 April 2025</i>
---	--

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = BMDC 8164.14.00 09/16
04/21/2025 11:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01524080A4
Standard Cyl #? 27
Standard Expiration? 03/05/2026
Oper No? 131856

Flow Cal. Date: 01/18/2010
Slope 732
Intercept -767769

IR Calibration Date: 04/29/2024
 3um 9um

0th Coef(*100):	-24811	-25942
1st Coef(*100):	260347	130533
2nd Coef(*100):	2582	1344
H2O adj (mg/l*10k):	289	383

***** Printer Test End *****



Operator Signature
TAYLOR WRAY

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = BMDC 8164.14.00 09/16
04/21/2025 11:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Gas	0.079	11:20
03 Room Air	0.000	11:20
04 Std. Gas	0.078	11:21
05 Room Air	0.000	11:21
06 Std. Gas	0.079	11:22
07 Room Air	0.000	11:22

Lot No = 01524080A4
Cyl No = 27
Exp Date = 03/05/2026
County = 08 Oper No. = 131856


Operator Signature
TAYLOR WRAY

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = BMDC 8164.14.00 09/16
04/21/2025 11:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	11:23
02 Room Air	0.000	11:24

*Invalid Test
Inhibited - RFI

Lot No = 01524080A4
Cyl No = 27
Exp Date = 03/05/2026
County = 08 Oper No. = 131856


Operator Signature
TAYLOR WRAY

Remarks:

Form 106-I8000