



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80-007093	Instrument Location BOPD
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000. **Fax. (523-7354)**
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): **Annual Inspection**
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature [Signature]	Date 4/24/2025
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Narehood	Certified Date 28April2025
---	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
04/24/2025 17:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN O PQRSTU VWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN O PQRSTU VWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 006
Standard Expiration? 10/05/2026
Oper No? 130538

Flow Cal. Date: 06/07/2019
Slope 638
Intercept -611758

IR Calibration Date: 12/16/2019
 3um 9um

0th Coef(*100):	-14610	-32356
1st Coef(*100):	264615	138933
2nd Coef(*100):	2992	1364
H2O adj(mg/l*10k):	369	621

***** Printer Test End *****


Operator Signature
TREVOR BERGERSON

Remarks:

Install after Annual Inspection.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
04/24/2025 17:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:43
02 Std. Gas	0.081	17:43
03 Room Air	0.000	17:44
04 Std. Gas	0.080	17:44
05 Room Air	0.000	17:45
06 Std. Gas	0.080	17:45
07 Room Air	0.000	17:46

Lot No = 25024080A3
Cyl No = 006
Exp Date = 10/05/2026
County = 06 Oper No. = 130538

 #1821

Operator Signature
TREVOR BERGERSON

Remarks:

Install after Annual Inspection

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
04/24/2025 17:46

Test	AC	Time
01 Room Air	RFI*	17:48
02 Room Air	0.000	17:48

*Invalid Test
Inhibited - RFI

Sub Name = RFI, RFI RFI
Sub DOB = 06/03/1988
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 25024080A3
Cyl No = 006
Expiration Date = 10/05/2026
County = 06 Oper No. = 130538

 #1821

Operator Signature
TREVOR BERGERSON

Remarks:

Install after Annual Inspection.

Form 106-I8000