



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6493 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 14323080A4 Cyl. # 14)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: WARD
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 790 psi Regulator: 750 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - High AC (Level 1,C)  
Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP16035  
Lot #: 202402C  
Exp. Date: 2/14/26  
☒ Results  $\pm 5\%$  AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
Known Value : 0.080 AC  
Gas Cylinder Lot #: 14323080A1  
Cylinder #: 69  
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)  
Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7352  
Lot #: 1CS8  
Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)  
Test 1 0.082 Test 4 0.082 Test 7 0.082  
Test 2 0.081 Test 5 0.081 Test 8 0.082  
Test 3 0.081 Test 6 0.081 Test 9 0.082  
Average 0.081  
☒ Results  $\pm 0.005$  AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]  
Inspector Signature

25 April 2025  
Date

[Signature]  
Reviewer

28 Apr 2025  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:11

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 09:11 |
| 02 Std. Sol. | 0.022 | 09:12 |
| 03 Room Air  | 0.000 | 09:13 |
| 04 Std. Sol. | 0.021 | 09:13 |
| 05 Room Air  | 0.000 | 09:14 |
| 06 Std. Sol. | 0.022 | 09:15 |
| 07 Room Air  | 0.000 | 09:15 |

08 Sim Temp = 34.0°C

Simul Ser No = MP5321  
Std Sol No = 202311D  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Low AC Check - 0.020AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:22

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 09:23 |
| 02 Std. Sol. | 0.294 | 09:24 |
| 03 Room Air  | 0.000 | 09:24 |
| 04 Std. Sol. | 0.295 | 09:25 |
| 05 Room Air  | 0.000 | 09:25 |
| 06 Std. Sol. | 0.295 | 09:26 |
| 07 Room Air  | 0.000 | 09:27 |

08 Sim Temp = 34.0°C

Simul Ser No = MP6035  
Std Sol No = 202402C  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:27

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 09:28 |
| 02 *Subject Test | INT*  | 09:29 |
| 03 Room Air      | 0.000 | 09:29 |

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A1  
Cyl No = 69  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:30

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Diagnostic    | OK    | 09:30 |
| 02 Room Air      | 0.000 | 09:31 |
| 03 *Subject Test | RFI*  | 09:31 |
| 04 Room Air      | 0.000 | 09:32 |

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A1  
Cyl No = 69  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:32

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:32 |
| 02 Std. Gas | 0.082 | 09:33 |
| 03 Room Air | 0.000 | 09:33 |
| 04 Std. Gas | 0.081 | 09:34 |
| 05 Room Air | 0.000 | 09:34 |
| 06 Std. Gas | 0.081 | 09:35 |
| 07 Room Air | 0.000 | 09:35 |

Lot No = 14323080A1  
Cyl No = 69  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks:

dry cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:36

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:36 |
| 02 Std. Gas | 0.082 | 09:37 |
| 03 Room Air | 0.000 | 09:37 |
| 04 Std. Gas | 0.081 | 09:38 |
| 05 Room Air | 0.000 | 09:38 |
| 06 Std. Gas | 0.081 | 09:39 |
| 07 Room Air | 0.000 | 09:39 |

Lot No = 14323080A1  
Cyl No = 69  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006493  
Location = TOXL      8164.14.00 09/16  
04/25/2025      09:39

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:40 |
| 02 Std. Gas | 0.082 | 09:40 |
| 03 Room Air | 0.000 | 09:41 |
| 04 Std. Gas | 0.082 | 09:41 |
| 05 Room Air | 0.000 | 09:42 |
| 06 Std. Gas | 0.082 | 09:42 |
| 07 Room Air | 0.000 | 09:43 |

Lot No = 14323080A1  
Cyl No = 69  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: DRY Cal check #7-9

Form 106-I8000