




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80 - 006676	Instrument Location MOSO
Reason for Install/Repair ☒ Install After Receiving From Crime Laboratory ☐ Install After Location Change ☐ Other (Specify) _____	


Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☐ 7. Repair and/or Maintenance Performed (if any): _____
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  282	Date 04/25/2025
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Narehood 	Certified Date 29April2025
--	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MOSO 8164.14.00 09/16
04/25/2025 17:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 3
Standard Expiration? 11/05/2025
Oper No? 131334

Flow Cal. Date: 08/22/2016
Slope 710
Intercept -713681

IR Calibration Date: 08/22/2016
3um 9um

0th Coef(*100):	-16667	-22327
1st Coef(*100):	259190	132291
2nd Coef(*100):	3393	1374
H2O adj(mg/l*10k):	611	546

***** Printer Test End *****


Operator Signature
JOSHUA GUDVANGEN

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MOSO 8164.14.00 09/16
04/25/2025 17:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:17
02 Std. Gas	0.083	17:17
03 Room Air	0.000	17:18
04 Std. Gas	0.082	17:18
05 Room Air	0.000	17:19
06 Std. Gas	0.082	17:19
07 Room Air	0.000	17:20

Lot No = 28423080A3

Cyl No = 3

Exp Date = 11/05/2025

County = 31

Oper No. = 131334


Operator Signature
JOSHUA GUDVANGEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MOSO 8164.14.00 09/16
04/25/2025 17:20

Test	AC	Time
01 Room Air	0.000	17:21
02 Reference	RFI*	17:22
03 Room Air	0.000	17:22

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST RFI TEST
Sub DOB = 01/01/1900
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 28423080A3
Cyl No = 3
Expiration Date = 11/05/2025
County = 31 Oper No. = 131334

 282

Operator Signature
JOSHUA GUDVANGEN

Remarks:

Form 106-I8000