



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 10687

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder ☒ Yes or No (If Yes, Lot # 14323080A4 Cyl. # 7)
 - Keys ☒ Yes or No
 - Power Cord ☒ Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: PLPD
5. ☒ Battery Check
 - Was the external battery pack replaced? Yes or ☒ No
6. ☒ O-Rings
 - Replaced Simulator O-Ring Yes or ☒ No
 - Replaced Breath Tube O-Ring ☒ Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
 - Display: 453 psi Regulator: 500 psi
 - Display and Regulator \pm 50 psi of each other ☒ Yes or No
 - Gas tank tare necessary? ☒ Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value \leq 0.03 AC: 0.020 AC
 - Sim. Ser #: MP5321
 - Lot #: 202311D
 - Exp. Date: 11/28/25
 - ☒ Results \pm 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results \pm 5% AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A1
Cylinder #: 71
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: ICS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.081 Test 4 0.082 Test 7 0.081
Test 2 0.082 Test 5 0.082 Test 8 0.081
Test 3 0.081 Test 6 0.082 Test 9 0.081
Average 0.081
☒ Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

Am. [Signature]
Inspector Signature

22 April 2025
Date

Janella Portscholler
Reviewer

24 Apr 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 7
Standard Expiration? 06/05/2025
Oper No? 133237

Flow Cal. Date: 10/05/2018
Slope 696
Intercept -609977

IR Calibration Date: 10/05/2018
 3um 9um

0th Coef(*100): -11478 -18771
1st Coef(*100): 255256 132135
2nd Coef(*100): 4432 1676
H2O adj(mg/l*10k): 604 575

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Sol.	0.020	14:37
03 Room Air	0.000	14:37
04 Std. Sol.	0.020	14:38
05 Room Air	0.000	14:38
06 Std. Sol.	0.021	14:39
07 Room Air	0.000	14:39

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check- 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:40


WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Sol.	0.294	14:42
03 Room Air	0.000	14:43
04 Std. Sol.	0.296	14:43
05 Room Air	0.000	14:44
06 Std. Sol.	0.296	14:45
07 Room Air	0.000	14:45

08 Sim Temp = 34.0°C

Simul Ser No = MP6035
Std Sol No = 202402C
County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:51

Test	AC	Time
01 Room Air	0.000	14:52
02 *Subject Test	INT*	14:52
03 Room Air	0.000	14:53

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 71
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:48

Test	AC	Time
01 Diagnostic	OK	14:49
02 Room Air	RFI*	14:49
03 Room Air	0.000	14:49

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 71
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.081	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.082	14:55
05 Room Air	0.000	14:56
06 Std. Gas	0.081	14:56
07 Room Air	0.000	14:56

Lot No = 14323080A1
Cyl No = 71
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Gas	0.082	14:58
03 Room Air	0.000	14:59
04 Std. Gas	0.082	14:59
05 Room Air	0.000	15:00
06 Std. Gas	0.082	15:00
07 Room Air	0.000	15:00

Lot No = 14323080A1
Cyl No = 71
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = TOXL 8164.14.00 09/16
04/22/2025 15:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Gas	0.081	15:02
03 Room Air	0.000	15:02
04 Std. Gas	0.081	15:03
05 Room Air	0.000	15:03
06 Std. Gas	0.081	15:03
07 Room Air	0.000	15:04

Lot No = 14323080A1
Cyl No = 71
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000