NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5943 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
   - Gas Cylinder — Yes or No (If Yes, Lot # 28423080A3 Cyl. # 42)
   - Keys — Yes or No
   - Power Cord — Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: NCSO
5. ☒ Battery Check
   - Was the external battery pack replaced? — Yes or No
6. ☒ O-Rings
   - Replaced Simulator O-Ring — Yes or No
   - Replaced Breath Tube O-Ring — Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
   - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
   - Display: 776 psi Regulator: 750 psi
   - Display and Regulator ± 50 psi of each other — Yes or No
   - Gas tank tare necessary? — Yes or No
   - If Yes, display readings after tare (Level 3,M,C,G):
   - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
   - Wet Calibration Check - Low AC (Level 1,C)
   - Known Value ≤ 0.03 AC: 0.020 AC
   - Sim. Ser #: MP5321
   - Lot #: 202311D
   - Exp. Date: 11/28/25
   - ☒ Results ± 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
   Wet Calibration Check - High AC (Level 1,C)
   Known Value ≥ 0.25 AC: 0.300 AC
   Sim. Ser #: MP6035
   Lot #: 202402C
   Exp. Date: 2/14/26
   ☒ Results ± 5% AC of known AC

3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
   Known Value : 0.080 AC
   Gas Cylinder Lot #: 1432308DA4
   Cylinder #: 6
   Exp. Date: 10/5/25

4. ☒ Interferent Check (Level 1,B)
   Known Value: 0.10 AC + 0.05% Acetone
   Sim. Ser #: DR7352
   Lot #: ICS 8
   Exp. Date: None
   ☒ Display reads "Interferent Detect"

5. ☒ RFI Check (CMS Mode)
   ☒ Display reads "RFI Detect"

6. ☒ Dry Calibration Check (Level 1,C)

Test 1 0.081	Test 4 0.081	Test 7 0.082
Test 2 0.081	Test 5 0.082	Test 8 0.081
Test 3 0.081	Test 6 0.081	Test 9 0.082
	Average 0.081	

   ☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. (Yes) or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

Inspector Signature — Date: 01May2025

Reviewer — Date: 01 May 2025

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = TOXL 8164.14.00 09/16
05/01/2025 09:10

********** Printer Test **********

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?

Current Instrument Setup

Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	28423080A3
Standard Cyl #?	42
Standard Expiration?	11/05/2025
Oper No?	133237
Flow Cal. Date:	06/20/2013
Slope	686
Intercept	-735246
IR Calibration Date:	08/26/2016

	3um	9um
0th Coef(*100):	-13589	-19547
1st Coef(*100):	284566	139614
2nd Coef(*100):	2717	1336
H2O adj(mg/l*10k):	381	404

********** Printer Test End **********

Operator Signature
ANNA NAREHOOD

Remarks: print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005943
Location = TOXL — 8164.14.00 09/16
05/01/2025 — 09:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:11
02 Std. Sol.	0.021	09:12
03 Room Air	0.000	09:12
04 Std. Sol.	0.021	09:13
05 Room Air	0.000	09:14
06 Std. Sol.	0.021	09:14
07 Room Air	0.000	09:15

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 — Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005943
Location = TOXL — 8164.14.00 09/16
05/01/2025 — 09:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:17
02 Std. Sol.	0.297	09:18
03 Room Air	0.000	09:18
04 Std. Sol.	0.298	09:19
05 Room Air	0.000	09:20
06 Std. Sol.	0.298	09:20
07 Room Air	0.000	09:21

08 Sim Temp = 34.0°C

Simul Ser No = MP6035
Std Sol No = 202402C
County = 08 — Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = TOXL 8164.14.00 09/16
05/01/2025 09:23

Test	AC	Time
01 Room Air	0.000	09:23
02 *Subject Test	INT*	09:24
03 Room Air	0.000	09:24

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 6
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = TOXL 8164.14.00 09/16
05/01/2025 09:25

Test	AC	Time
01 Diagnostic	OK	09:25
02 Room Air	RFI*	09:26
03 Room Air	0.000	09:26

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 6
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005943
Location = TOXL — 8164.14.00 09/16
05/01/2025 — 09:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:34
02 Std. Gas	0.081	09:34
03 Room Air	0.000	09:35
04 Std. Gas	0.081	09:35
05 Room Air	0.000	09:36
06 Std. Gas	0.081	09:36
07 Room Air	0.000	09:37

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 — Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005943
Location = TOXL — 8164.14.00 09/16
05/01/2025 — 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Gas	0.081	09:38
03 Room Air	0.000	09:39
04 Std. Gas	0.082	09:39
05 Room Air	0.000	09:40
06 Std. Gas	0.081	09:40
07 Room Air	0.000	09:40

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 — Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer 8000 — Alcohol Analyzer
North Dakota Model — SN 80-005943
Location = TOXL — 8164.14.00 09/16
05/01/2025 — 09:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:42
02 Std. Gas	0.082	09:42
03 Room Air	0.000	09:43
04 Std. Gas	0.081	09:43
05 Room Air	0.000	09:44
06 Std. Gas	0.082	09:44
07 Room Air	0.000	09:45

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 — Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #7-9

Form 106-I8000