



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3057 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: TOXL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 753 psi Regulator: 725 psi
Display and Regulator $\pm$ 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value $\leq$ 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value ≥ 0.25 AC: 0.300 AC
 Sim. Ser #: MP6035
 Lot #: 202402C
 Exp. Date: 2114126
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 6
 Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7352
 Lot #: 1CS8
 Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.081</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>
Average <u>0.080</u>		

☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or (No)

If No, state reason(s) why: Classroom only instrument due to a flow sensor issues from prior Annual Inspections.

If Yes, change location code back to A.4 unless A.4 is TOXL. ☐

[Signature]
Inspector Signature

02 May 2025
Date

Janelle Pritschler
Reviewer

05 May 2025
Date

CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-003057
Location = TOXL	8164.14.00 09/16
05/02/2025	12:22

```
abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?
```

```
abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*() +?
```

Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	02621080A1
Standard Cyl #?	6
Standard Expiration?	03/05/2023
Oper No?	133237

```
Flow Cal. Date:      06/14/2021
Slope                516
Intercept            495821
```

IR Calibration Date: 06/14/2021
3um 9um

0th Coef(*100):	-20706	-12354
1st Coef(*100):	264169	140515
2nd Coef(*100):	1074	859
H2O adj (mg/l*10k):	351	239

Am 1.1.2020

Operator Signature
ANNA NAREHOOD

Remarks: print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Sol.	0.021	12:37
03 Room Air	0.000	12:38
04 Std. Sol.	0.020	12:38
05 Room Air	0.000	12:39
06 Std. Sol.	0.020	12:40
07 Room Air	0.000	12:40

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Sol.	0.291	12:44
03 Room Air	0.000	12:45
04 Std. Sol.	0.292	12:45
05 Room Air	0.000	12:46
06 Std. Sol.	0.293	12:47
07 Room Air	0.000	12:47

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:50

Test	AC	Time
01 Room Air	0.000	12:51
02 *Subject Test	INT*	12:51
03 Room Air	0.000	12:52

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 6
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:52

Test	AC	Time
01 Diagnostic	OK	12:53
02 Room Air	0.000	12:53
03 *Subject Test	RFI*	12:54
04 Room Air	0.000	12:54

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 6
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:55
02 Std. Gas	0.081	12:56
03 Room Air	0.000	12:56
04 Std. Gas	0.081	12:56
05 Room Air	0.000	12:57
06 Std. Gas	0.081	12:57
07 Room Air	0.000	12:58

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Drycal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 12:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Gas	0.081	12:59
03 Room Air	0.000	13:00
04 Std. Gas	0.081	13:00
05 Room Air	0.000	13:01
06 Std. Gas	0.080	13:01
07 Room Air	0.000	13:02

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Drycal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.14.00 09/16
05/02/2025 13:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:02
02 Std. Gas	0.081	13:03
03 Room Air	0.000	13:03
04 Std. Gas	0.080	13:04
05 Room Air	0.000	13:04
06 Std. Gas	0.081	13:04
07 Room Air	0.000	13:05

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: dry cal check #7-9

Form 106-I8000