



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Cody D. Meadows</i>		
Location <i>Washburn</i>	Intoxilyzer® Serial Number <i>80-006674</i>	
Gas Lot Number <i>08425080A1</i>	Gas Cylinder Number <i>6</i>	Gas Expiration Date <i>5/5/2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.081* AC
 - C. *0.081* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>5/2/25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>05May2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = WASH 8164.14.00 09/16
05/02/2025 11:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:28
02 Std. Gas	0.081	11:29
03 Room Air	0.000	11:29
04 Std. Gas	0.081	11:30
05 Room Air	0.000	11:30
06 Std. Gas	0.081	11:31
07 Room Air	0.000	11:31

Lot No = 08425080A1
Cyl No = 6
Exp Date = 05/05/2027
County = 28 Oper No. = 100724



Operator Signature
CODY MEADOWS

Remarks:

Form 106-I8000