



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Ryan Haskell		
Location WFPD		Intoxilyzer® Serial Number 80-006684
Gas Lot Number 25024080A3	Gas Cylinder Number 11	Gas Expiration Date 10/05/2026

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0.081 AC
B. 0.082 AC
C. 0.082 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature 	Date 5/2/2025
Reviewed By (Crime Laboratory Use Only) Anna Narehood	Date 02May2025


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
05/02/2025 08:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Gas	0.081	08:47
03 Room Air	0.000	08:48
04 Std. Gas	0.082	08:48
05 Room Air	0.000	08:49
06 Std. Gas	0.082	08:49
07 Room Air	0.000	08:50

Lot No = 25024080A3
Cyl No = 11
Exp Date = 10/05/2026
County = 09 Oper No. = 121019


Operator Signature
RYAN HASKELL

Remarks:

Form 106-I8000