

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number <b>80-006503</b>	Instrument Location <b>MINOT PD 515 2ND AVE SW</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	


## Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
  - ☒ A. Print Test (Level 1, Function P).
  - ☒ B. ACA Test (Level 1, Function C).
  - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>05/02/25</b>
---	-------------------------

## Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Anna Narehood</b> 	Certified Date <b>06May2025</b>
--	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006503  
Location = MIPD      8164.14.00 09/16  
05/02/2025      14:16

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      08425080A1  
Standard Cyl #?      10  
Standard Expiration?      05/05/2027  
Oper No?      132576

Flow Cal. Date:      08/20/2015  
Slope      663  
Intercept      -633300

IR Calibration Date:      06/12/2017  
                                 3um      9um

0th Coef(*100):	-6741	-15632
1st Coef(*100):	256583	132194
2nd Coef(*100):	3028	1131
H2O adj(mg/l*10k):	354	358

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ZACKERY MYERS

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006503  
Location = MIPD      8164.14.00 09/16  
05/02/2025      14:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:17
02 Std. Gas	0.082	14:17
03 Room Air	0.000	14:18
04 Std. Gas	0.081	14:18
05 Room Air	0.000	14:19
06 Std. Gas	0.082	14:19
07 Room Air	0.000	14:19

Lot No = 08425080A1  
Cyl No = 10  
Exp Date = 05/05/2027  
County = 51      Oper No. = 132576



Operator Signature  
ZACKERY MYERS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006503  
Location = MIPD      8164.14.00 09/16  
05/02/2025      14:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	14:21
02 Room Air	0.000	14:22

\*Invalid Test  
Inhibited - RFI

Lot No = 08425080A1  
Cyl No = 10  
Exp Date = 05/05/2027  
County = 51      Oper No. = 132576

  
\_\_\_\_\_  
Operator Signature  
ZACKERY MYERS

Remarks:

Form 106-I8000