



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3069

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: TOXL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 372 psi Regulator: 350 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5321 202311D aen 5/5/25
Lot #: 412-aen 5/5/25 202311D
Exp. Date: 11/28/25
☒ Results \pm 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results \pm 5% AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A1
Cylinder #: 69
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: 1CS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.080 Test 4 0.080 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.079
Test 3 0.079 Test 6 0.080 Test 9 0.081
Average 0.079
☒ Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: per prior annual inspections, this instrument is classroom only due to no battery status when unplugged.

If Yes, change location code back to A.4 unless A.4 is TOXL. ☐

Jim Miller
Inspector Signature

05 May 2025
Date

Janelle Putschler
Reviewer

05 May 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:20

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 23
Standard Expiration? 03/05/2023
Oper No? 133237

Flow Cal. Date: 03/30/2022
Slope 662
Intercept -676090

IR Calibration Date: 03/30/2022
 3um 9um

0th Coef(*100): -20151 -12325
1st Coef(*100): 267293 130383
2nd Coef(*100): 4119 1500
H2O adj(mg/l*10k): 550 327

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Sol.	0.019	11:23
03 Room Air	0.000	11:23
04 Std. Sol.	0.019	11:24
05 Room Air	0.000	11:24
06 Std. Sol.	0.019	11:25
07 Room Air	0.000	11:26

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Low AC Check -0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Sol.	0.292	11:28
03 Room Air	0.000	11:29
04 Std. Sol.	0.294	11:29
05 Room Air	0.000	11:30
06 Std. Sol.	0.295	11:31
07 Room Air	0.000	11:31

08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:32

Test	AC	Time
01 Room Air	0.000	11:33
02 *Subject Test	INT*	11:34
03 Room Air	0.000	11:35

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 69
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:35

Test	AC	Time
01 Diagnostic	OK	11:36
02 Room Air	0.000	11:36
03 *Subject Test	RFI*	11:36
04 Room Air	0.000	11:37

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 69
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Gas	0.080	11:38
03 Room Air	0.000	11:39
04 Std. Gas	0.080	11:39
05 Room Air	0.000	11:39
06 Std. Gas	0.079	11:40
07 Room Air	0.000	11:40

Lot No = 14323080A1
Cyl No = 69
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *dry cal check #1-3*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.080	11:42
03 Room Air	0.000	11:42
04 Std. Gas	0.080	11:43
05 Room Air	0.000	11:43
06 Std. Gas	0.080	11:43
07 Room Air	0.000	11:44

Lot No = 14323080A1
Cyl No = 69
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: DRY CAL CHECK #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.14.00 09/16
05/05/2025 11:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:46
02 Std. Gas	0.080	11:46
03 Room Air	0.000	11:47
04 Std. Gas	0.079	11:47
05 Room Air	0.000	11:48
06 Std. Gas	0.081	11:48
07 Room Air	0.000	11:48

Lot No = 14323080A1
Cyl No = 69
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #7-9

Form 106-I8000