



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4943

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 14323080A1 Cyl. # 16)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: BCCC
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 972 psi Regulator: 975 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - High AC (Level 1,C)  
Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP6035  
Lot #: 202402C  
Exp. Date: 02/14/26  
☒ Results  $\pm$  5% AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
Known Value : 0.080 AC  
Gas Cylinder Lot #: 14323080A4  
Cylinder #: 6  
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)  
Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7352  
Lot #: 1CS8  
Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)  
Test 1 0.081 Test 4 0.081 Test 7 0.082  
Test 2 0.082 Test 5 0.081 Test 8 0.081  
Test 3 0.081 Test 6 0.082 Test 9 0.081  
Average 0.081  
☒ Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]  
Inspector Signature

07 May 2025  
Date

[Signature]  
Reviewer

08 May 2025  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      11:55

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      14323080A1  
Standard Cyl #?      16  
Standard Expiration?      06/05/2025  
Oper No?      133237

Flow Cal. Date:      04/06/2020  
Slope      679  
Intercept      -694345

IR Calibration Date:      04/06/2020  
   3um      9um

0th Coef(*100):	-23601	-21377
1st Coef(*100):	255715	132279
2nd Coef(*100):	1319	1274
H2O adj (mg/l*10k):	424	419

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
ANNA NAREHOOD

Remarks:      print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      11:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Sol.	0.020	12:00
03 Room Air	0.000	12:01
04 Std. Sol.	0.020	12:01
05 Room Air	0.000	12:02
06 Std. Sol.	0.020	12:03
07 Room Air	0.000	12:03

08 Sim Temp = 34.0°C

Simul Ser No = MP5321  
Std Sol No = 202311D  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: LOW AC Check - 0.020AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Sol.	0.292	12:06
03 Room Air	0.000	12:06
04 Std. Sol.	0.294	12:07
05 Room Air	0.000	12:07
06 Std. Sol.	0.294	12:08
07 Room Air	0.000	12:09


08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:10

Test	AC	Time
01 Room Air	0.000	12:11
02 *Subject Test	INT*	12:12
03 Room Air	0.000	12:12

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 6  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*  
*0.10AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:13

Test	AC	Time
01 Diagnostic	OK	12:14
02 Room Air	RFI*	12:14
03 Room Air	0.000	12:14

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 6  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Gas	0.081	12:16
03 Room Air	0.000	12:16
04 Std. Gas	0.082	12:17
05 Room Air	0.000	12:17
06 Std. Gas	0.081	12:18
07 Room Air	0.000	12:18

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:19
02 Std. Gas	0.081	12:20
03 Room Air	0.000	12:20
04 Std. Gas	0.081	12:20
05 Room Air	0.000	12:21
06 Std. Gas	0.082	12:21
07 Room Air	0.000	12:22

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004943  
Location = TOXL      8164.14.00 09/16  
05/07/2025      12:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:25
02 Std. Gas	0.082	12:25
03 Room Air	0.000	12:26
04 Std. Gas	0.081	12:26
05 Room Air	0.000	12:27
06 Std. Gas	0.081	12:27
07 Room Air	0.000	12:28

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: DRY CAL CHECK #7-9

Form 106-I8000