



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7095 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # 1432308041 Cyl. # 03)  
Keys Yes or No  
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: OKPD
5. ☒ Battery Check  
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).  
Display: 075 psi Regulator: 050 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5321  
Lot #: 202311D  
Exp. Date: 11/28/25  
☒ Results  $\pm$  0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq 0.25$  AC: 0.300 AC  
 Sim. Ser #: MP6035  
 Lot #: 202402C  
 Exp. Date: 2/14/26  
☒ Results  $\pm 5\%$  AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 4323080A4  
 Cylinder #: 6 ~~415~~ den 517125  
 Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7352  
 Lot #: 1CS8  
 Exp. Date: None  
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)  
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)  

|                      |                     |                     |
|----------------------|---------------------|---------------------|
| Test 1 <u>0.081</u>  | Test 4 <u>0.080</u> | Test 7 <u>0.081</u> |
| Test 2 <u>0.081</u>  | Test 5 <u>0.081</u> | Test 8 <u>0.081</u> |
| Test 3 <u>0.081</u>  | Test 6 <u>0.081</u> | Test 9 <u>0.081</u> |
| Average <u>0.080</u> |                     |                     |

☒ Results  $\pm 0.005$  AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

An Mill  
 Inspector Signature

07 May 2025  
 Date

Jamille Pertschiller  
 Reviewer

08 May 2025  
 Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      09:17

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      14323080A1  
Standard Cyl #?      63  
Standard Expiration?      06/05/2025  
Oper No?      133237

Flow Cal. Date:      09/27/2018  
Slope      683  
Intercept      -637452

IR Calibration Date:      06/21/2023  
                                         3um      9um

|                    |        |        |
|--------------------|--------|--------|
| 0th Coef(*100):    | -38728 | -15234 |
| 1st Coef(*100):    | 250485 | 139595 |
| 2nd Coef(*100):    | 2553   | 1269   |
| H2O adj(mg/l*10k): | 921    | 383    |

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      09:50

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 09:51 |
| 02 Std. Sol. | 0.021 | 09:52 |
| 03 Room Air  | 0.000 | 09:52 |
| 04 Std. Sol. | 0.020 | 09:53 |
| 05 Room Air  | 0.000 | 09:53 |
| 06 Std. Sol. | 0.021 | 09:54 |
| 07 Room Air  | 0.000 | 09:55 |

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Low AC check -0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      09:58

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 09:58 |
| 02 Std. Sol. | 0.300 | 09:59 |
| 03 Room Air  | 0.000 | 09:59 |
| 04 Std. Sol. | 0.300 | 10:00 |
| 05 Room Air  | 0.000 | 10:01 |
| 06 Std. Sol. | 0.300 | 10:01 |
| 07 Room Air  | 0.000 | 10:02 |

08 Sim Temp = 34.0°C

Simul Ser No = MP6035  
Std Sol No = 202402C  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      10:04

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 10:04 |
| 02 *Subject Test | INT*  | 10:05 |
| 03 Room Air      | 0.000 | 10:06 |

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 6  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Interferent check  
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      10:08

| Test          | AC    | Time  |
|---------------|-------|-------|
| 01 Diagnostic | OK    | 10:09 |
| 02 Room Air   | RFI*  | 10:09 |
| 03 Room Air   | 0.000 | 10:10 |

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 6  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      10:10

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:11 |
| 02 Std. Gas | 0.081 | 10:11 |
| 03 Room Air | 0.000 | 10:11 |
| 04 Std. Gas | 0.081 | 10:12 |
| 05 Room Air | 0.000 | 10:12 |
| 06 Std. Gas | 0.081 | 10:13 |
| 07 Room Air | 0.000 | 10:13 |

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      10:15

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:16 |
| 02 Std. Gas | 0.080 | 10:16 |
| 03 Room Air | 0.000 | 10:17 |
| 04 Std. Gas | 0.081 | 10:17 |
| 05 Room Air | 0.000 | 10:18 |
| 06 Std. Gas | 0.081 | 10:18 |
| 07 Room Air | 0.000 | 10:18 |

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = TOXL      8164.16.00 09/18  
05/07/2025      10:19

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:19 |
| 02 Std. Gas | 0.081 | 10:20 |
| 03 Room Air | 0.000 | 10:20 |
| 04 Std. Gas | 0.081 | 10:21 |
| 05 Room Air | 0.000 | 10:21 |
| 06 Std. Gas | 0.081 | 10:21 |
| 07 Room Air | 0.000 | 10:22 |

Lot No = 14323080A4  
Cyl No = 6  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000