



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>William Earl</i>		
Location <i>DEVL</i>		Intoxilyzer® Serial Number <i>80-007096</i>
Gas Lot Number <i>08425080A1</i>	Gas Cylinder Number <i>018</i>	Gas Expiration Date <i>5/5/2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.080 AC
 - B. 0.077 AC
 - C. 0.079 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>William Earl</i>	Date <i>9/7/2025</i>
Reviewed By (Crime Laboratory Use Only) Anna Narehood <i>Anna Narehood</i>	Date 08May2025


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007096
Location = DEVL 8164.16.00 09/18
05/07/2025 10:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Gas	0.080	10:25
03 Room Air	0.000	10:25
04 Std. Gas	0.079	10:26
05 Room Air	0.000	10:26
06 Std. Gas	0.079	10:27
07 Room Air	0.000	10:27

Lot No = 08425080A1
Cyl No = 018
Exp Date = 05/05/2027
County = 36 Oper No. = 132008


Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000