



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>William Earl</i>		
Location <i>DEVL</i>		Intoxilyzer® Serial Number <i>80-004944</i>
Gas Lot Number <i>25024080A3</i>	Gas Cylinder Number <i>004</i>	Gas Expiration Date <i>10/5/2024</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.082 AC
 - B. 0.082 AC
 - C. 0.082 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>William Earl</i>	Date <i>5/2/2025</i>
Reviewed By (Crime Laboratory Use Only) Anna Narehood <i>Anna Narehood</i>	Date 08May2025


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = DEVL 8164.14.00 09/16
05/07/2025 13:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:23
02 Std. Gas	0.082	13:23
03 Room Air	0.000	13:23
04 Std. Gas	0.082	13:24
05 Room Air	0.000	13:24
06 Std. Gas	0.082	13:25
07 Room Air	0.000	13:25

Lot No = 25024080A3
Cyl No = 004
Exp Date = 10/05/2026
County = 36 Oper No. = 132008


Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000