



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 10689 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # 14323080A4 Cyl. # 40)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: DKPD
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 334 psi Regulator: 300 psi
Display and Regulator $\pm$ 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value $\leq$ 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP16035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 28423080A3
Cylinder #: 5
Exp. Date: 11/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: ICS8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.077 Test 4 0.078 Test 7 0.077
Test 2 0.077 Test 5 0.078 Test 8 0.077
Test 3 0.077 Test 6 0.077 Test 9 0.077
Average 0.077
☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

13 May 2025
Date

Janelle Putscheller
Reviewer

14 May 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.020
Standard Solution #? 202311D
Sim Serial #? MP5321
Oper No? 133237

Flow Cal. Date: 05/29/2020
Slope 643
Intercept -576193

IR Calibration Date: 05/29/2020
 3um 9um

0th Coef(*100):	-25870	-28277
1st Coef(*100):	287704	139899
2nd Coef(*100):	26	744
H2O adj(mg/l*10k):	293	328

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:26
02 Std. Sol.	0.021	12:27
03 Room Air	0.000	12:27
04 Std. Sol.	0.022	12:28
05 Room Air	0.000	12:29
06 Std. Sol.	0.021	12:29
07 Room Air	0.000	12:30

08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Low AC Check - 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Sol.	0.294	12:33
03 Room Air	0.000	12:34
04 Std. Sol.	0.296	12:34
05 Room Air	0.000	12:35
06 Std. Sol.	0.296	12:36
07 Room Air	0.000	12:36


08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Al Check -0.300Al

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:37

Test	AC	Time
01 Room Air	0.000	12:38
02 *Subject Test	INT*	12:39
03 Room Air	0.000	12:39

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CJ
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 5
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Check
0.10AL + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:40

Test	AC	Time
01 Diagnostic	OK	12:41
02 Room Air	0.000	12:41
03 *Subject Test	RFI*	12:41
04 Room Air	0.000	12:42

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 5
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Gas	0.077	12:44
03 Room Air	0.000	12:44
04 Std. Gas	0.077	12:45
05 Room Air	0.000	12:45
06 Std. Gas	0.077	12:46
07 Room Air	0.000	12:46

Lot No = 28423080A3
Cyl No = 5
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:47
02 Std. Gas	0.078	12:48
03 Room Air	0.000	12:48
04 Std. Gas	0.078	12:48
05 Room Air	0.000	12:49
06 Std. Gas	0.077	12:49
07 Room Air	0.000	12:50

Lot No = 28423080A3
Cyl No = 5
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check # 4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
05/13/2025 12:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Gas	0.077	12:52
03 Room Air	0.000	12:52
04 Std. Gas	0.077	12:52
05 Room Air	0.000	12:53
06 Std. Gas	0.077	12:53
07 Room Air	0.000	12:54

Lot No = 28423080A3
Cyl No = 5
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check # 7-9

Form 106-I8000