



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 10077
0607aen 5/15/25 Inspection Location: TDXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 14323080A1 Cyl. # 64)
 - Keys Yes or No
 - Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: GLDN
5. ☒ Battery Check
 - Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
 - Display: 317 psi Regulator: 300 psi
 - Display and Regulator \pm 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value \leq 0.03 AC: 0.020 AC
 - Sim. Ser #: MP5321
 - Lot #: 202311D
 - Exp. Date: 11/28/25
 - ☒ Results \pm 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP6035
Lot #: 202402C
Exp. Date: 2/14/26
☒ Results $\pm 5\%$ AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A4
Cylinder #: 42
Exp. Date: 05/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: 1C58
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.082 Test 4 0.082 Test 7 0.082
Test 2 0.082 Test 5 0.082 Test 8 0.082
Test 3 0.082 Test 6 0.082 Test 9 0.082
Average 0.082
☒ Results ± 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

15 May 2025
Date

[Signature]
Reviewer

17 May 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A1
Standard Cyl #? 64
Standard Expiration? 06/05/2025
Oper No? 133237

Flow Cal. Date: 05/22/2020
Slope 678
Intercept -711846

IR Calibration Date: 05/22/2020
3um 9um

0th Coef(*100):	-22305	-20910
1st Coef(*100):	281966	134470
2nd Coef(*100):	1570	1173
H2O adj(mg/l*10k):	336	351

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Phnt Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:09
02 Std. Sol.	0.022	09:10
03 Room Air	0.000	09:10
04 Std. Sol.	0.022	09:11
05 Room Air	0.000	09:12
06 Std. Sol.	0.021	09:12
07 Room Air	0.000	09:13

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:14
02 Std. Sol.	0.296	09:15
03 Room Air	0.000	09:16
04 Std. Sol.	0.297	09:17
05 Room Air	0.000	09:17
06 Std. Sol.	0.299	09:18
07 Room Air	0.000	09:18


08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check -0.300M

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:20

Test	AC	Time
01 Room Air	0.000	09:20
02 *Subject Test	INT*	09:21
03 Room Air	0.000	09:22

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 42
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Chem
0.10% Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:23

Test	AC	Time
01 Diagnostic	OK	09:24
02 Room Air	0.000	09:24
03 *Subject Test	RFI*	09:24
04 Room Air	0.000	09:25

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 42
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:26
02 Std. Gas	0.082	09:26
03 Room Air	0.000	09:27
04 Std. Gas	0.082	09:27
05 Room Air	0.000	09:28
06 Std. Gas	0.082	09:28
07 Room Air	0.000	09:29

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237


Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:30
02 Std. Gas	0.082	09:30
03 Room Air	0.000	09:31
04 Std. Gas	0.082	09:31
05 Room Air	0.000	09:32
06 Std. Gas	0.082	09:32
07 Room Air	0.000	09:33

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/15/2025 09:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:34
02 Std. Gas	0.082	09:34
03 Room Air	0.000	09:35
04 Std. Gas	0.082	09:35
05 Room Air	0.000	09:36
06 Std. Gas	0.082	09:36
07 Room Air	0.000	09:37

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Dry cal check #7-9

Form 106-I8000