



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Czapiewski, Rick</i>		
Location <i>FINL</i>	Intoxilyzer® Serial Number <i>80-005946</i>	
Gas Lot Number <i>08425080A1</i>	Gas Cylinder Number <i>8</i>	Gas Expiration Date <i>5/5/2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. *082* AC
 - B. 0. *080* AC
 - C. 0. *079* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Rick Czapiewski</i>	Date <i>5/9/2025</i>
Reviewed By (Crime Laboratory Use Only) Anna Narehood <i>Anna Narehood</i>	Date <i>15May2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = FINL 8164.14.00 09/16
05/09/2025 14:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:30
02 Std. Gas	0.082	14:31
03 Room Air	0.000	14:31
04 Std. Gas	0.080	14:31
05 Room Air	0.000	14:32
06 Std. Gas	0.079	14:32
07 Room Air	0.000	14:33

Lot No = 08425080A1
Cyl No = 008
Exp Date = 05/05/2027
County = 46 Oper No. = 130363


Operator Signature
RICK CZAPIEWSKI

Remarks:

Form 106-I8000