

**ETHANOL GAS STANDARD CYLINDER REPORT**

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Austin Bari</i>		
Location <i>HILL</i>		Intoxilyzer® Serial Number <i>80-006679</i>
Gas Lot Number <i>08425080A1</i>	Gas Cylinder Number <i>15</i>	Gas Expiration Date <i>5-5-2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.081* AC
B. *0.092* AC
C. *0.081* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>Austin Bari</i>	Date <i>5-17-25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>21May2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = HILL 8164.14.00 09/16
05/17/2025 16:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:52
02 Std. Gas	0.081	16:53
03 Room Air	0.000	16:53
04 Std. Gas	0.082	16:53
05 Room Air	0.000	16:54
06 Std. Gas	0.081	16:54
07 Room Air	0.000	16:55

Lot No = 08425080A1

Cyl No = 15

Exp Date = 05/05/2027

County = 49

Oper No. = 131630



Operator Signature
AUSTIN BARIL

Remarks:

Form 106-I8000