



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

|                                                              |                                                |                                          |
|--------------------------------------------------------------|------------------------------------------------|------------------------------------------|
| Chemical Test Operator Name (Print)<br><i>Muckle, Dey, J</i> |                                                |                                          |
| Location<br><i>Golden Valley County Sheriff's Office</i>     | Intoxilyzer® Serial Number<br><i>80-006667</i> |                                          |
| Gas Lot Number<br><i>08425080 A1</i>                         | Gas Cylinder Number<br><i>027</i>              | Gas Expiration Date<br><i>05/05/2027</i> |

## Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.081* AC
  - C. *0.082* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

|                                                                 |                           |
|-----------------------------------------------------------------|---------------------------|
| Chemical Test Operator Signature<br><i>[Signature]</i>          | Date<br><i>05/19/2025</i> |
| Reviewed By (Crime Laboratory Use Only)<br><i>Anna Narehood</i> | Date<br>20May2025         |

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = GLDN      8164.14.00 09/16  
05/19/2025      15:58

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:59 |
| 02 Std. Gas | 0.081 | 15:59 |
| 03 Room Air | 0.000 | 16:00 |
| 04 Std. Gas | 0.081 | 16:00 |
| 05 Room Air | 0.000 | 16:01 |
| 06 Std. Gas | 0.082 | 16:01 |
| 07 Room Air | 0.000 | 16:02 |


Lot No = 8425080A1

Cyl No = 27

Exp Date = 05/05/2027

County = 17

Oper No. = 130349

  
Operator Signature  
DEY MUCKLE

Remarks:

Form 106-I8000