



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Boerke, Danielle</i>		
Location <i>CCSO COOP</i>		Intoxilyzer® Serial Number <i>80-004956</i>
Gas Lot Number <i>084250 80A1</i>	Gas Cylinder Number <i>5</i>	Gas Expiration Date <i>05/05/2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.082* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>5/7/25</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Narehood</i>	Date <i>10 June 2025</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = COOP 8164.14.00 09/16
05/07/2025 13:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:58
02 Std. Gas	0.082	13:58
03 Room Air	0.000	13:59
04 Std. Gas	0.082	13:59
05 Room Air	0.000	14:00
06 Std. Gas	0.082	14:00
07 Room Air	0.000	14:01

Lot No = 08425080A1
Cyl No = 5
Exp Date = 05/05/2027
County = 20 Oper No. = 100754


Operator Signature
DANIELLE BJORLIE

Remarks:

Form 106-I8000