



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7087 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
Keys Yes or No
Power Cord Yes or No
2. ☐ Download Data
3. ☐ Upload Operator File
4. ☒ Current Location Code: TOXL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

unable to direct connect to instrument

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 499 psi Regulator: 500 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5321
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results \pm 0.005 of known AC


2. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - High AC (Level 1,C)
Known Value \geq 0.25 AC: 0.400 AC
Sim. Ser #: MP5318
Lot #: 202312D
Exp. Date: 12/7/25
☒ Results \pm 5% AC of known AC
3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).
Known Value : 0.080 AC
Gas Cylinder Lot #: 14323080A1
Cylinder #: 21
Exp. Date: 6/5/25
4. ☒ Interferent Check (Level 1,B)
Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7352
Lot #: ICS 8
Exp. Date: None
☒ Display reads "Interferent Detect"
5. ☒ RFI Check (CMS Mode)
☒ Display reads "RFI Detect"
6. ☒ Dry Calibration Check (Level 1,C)
Test 1 0.080 Test 4 0.080 Test 7 0.080
Test 2 0.080 Test 5 0.079 Test 8 0.080
Test 3 0.080 Test 6 0.080 Test 9 0.079
Average 0.079
☒ Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA


Instrument is acceptable to be used in the field. Yes or (No)

If No, state reason(s) why: unable to direct connect to instrument so operator file cannot be updated.

If Yes, change location code back to A.4 unless A.4 is TOXL. ☐


Inspector Signature

30 May 2025
Date


Reviewer

02 June 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 13:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 33
Standard Expiration? 04/05/2022
Oper No? 133237

Flow Cal. Date: 07/02/2019
Slope 721
Intercept -792109

IR Calibration Date: 07/02/2019
 3um 9um

0th Coef(*100):	-45706	-18865
1st Coef(*100):	253208	131287
2nd Coef(*100):	2874	1457
H2O adj(mg/l*10k):	1171	483

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Sol.	0.018	14:02
03 Room Air	0.000	14:02
04 Std. Sol.	0.018	14:03
05 Room Air	0.000	14:03
06 Std. Sol.	0.018	14:04
07 Room Air	0.000	14:05

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC Check -0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Sol.	0.395	14:10
03 Room Air	0.000	14:11
04 Std. Sol.	0.394	14:11
05 Room Air	0.000	14:12
06 Std. Sol.	0.393	14:13
07 Room Air	0.000	14:13

08 Sim Temp = 34.0°C

Simul Ser No = MP5318
Std Sol No = 202312D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.400 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:16

Test	AC	Time
01 Room Air	0.000	14:17
02 *Subject Test	INT*	14:17
03 Room Air	0.000	14:18

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 21
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:21

Test	AC	Time
01 Diagnostic	OK	14:21
02 Room Air	0.000	14:22
03 *Subject Test	RFI*	14:22
04 Room Air	0.000	14:23

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A1
Cyl No = 21
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:24
02 Std. Gas	0.080	14:24
03 Room Air	0.000	14:25
04 Std. Gas	0.080	14:25
05 Room Air	0.000	14:25
06 Std. Gas	0.080	14:26
07 Room Air	0.000	14:26

Lot No = 14323080A1
Cyl No = 21
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:27
02 Std. Gas	0.080	14:28
03 Room Air	0.000	14:28
04 Std. Gas	0.079	14:29
05 Room Air	0.000	14:29
06 Std. Gas	0.080	14:29
07 Room Air	0.000	14:30

Lot No = 14323080A1
Cyl No = 21
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: DRY cal check
#4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.16.00 09/18
05/30/2025 14:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:31
02 Std. Gas	0.080	14:31
03 Room Air	0.000	14:32
04 Std. Gas	0.080	14:32
05 Room Air	0.000	14:33
06 Std. Gas	0.079	14:33
07 Room Air	0.000	14:34

Lot No = 14323080A1
Cyl No = 21
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *dry cal check*
7-9

Form 106-I8000