



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7086 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # 25024080A3 Cyl. # 53)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: LSB
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 995 psi Regulator: 1000 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5324
Lot #: 202311D
Exp. Date: 11/28/25
☒ Results \pm 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP6035

Lot #: 202402C

Exp. Date: 2/14/26

☒ Results \pm 5% AC of known AC

3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 14323080A4

Cylinder #: 42

Exp. Date: 6/5/25

4. ☒ Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS8

Exp. Date: None

☒ Display reads "Interferent Detect"

5. ☒ RFI Check (CMS Mode)

☒ Display reads "RFI Detect"

6. ☒ Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.080 Test 7 0.080

Test 2 0.080 Test 5 0.080 Test 8 0.079

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.079

☒ Results \pm 0.005 AC of known AC

Post calibration adjust.

0.400 AC

Sim ser #: MP5318

Lot # 2023120

Exp. date: 12/7/25

☒ Results \pm 5% AC of known AC

D. Remarks/Maintenance: Two calibration dry checks were originally performed and it was determined that a calibration adjustment was necessary. The values in #6 are post calibration adjustment.

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☐ NIA. Instrument Swapped.

[Signature]
Inspector Signature

04 June 2025
Date

[Signature]
Reviewer

05 June 2025
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 25024080A3
Standard Cyl #? 053
Standard Expiration? 10/05/2026
Oper No? 133237

Flow Cal. Date: 10/16/2018
Slope 688
Intercept -653049

IR Calibration Date: 10/16/2018
3um 9um

0th Coef(*100):	-47466	-23263
1st Coef(*100):	233222	130299
2nd Coef(*100):	3639	1618
H2O adj(mg/l*10k):	1280	608

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Sol.	0.019	10:24
03 Room Air	0.000	10:24
04 Std. Sol.	0.019	10:25
05 Room Air	0.000	10:26
06 Std. Sol.	0.019	10:26
07 Room Air	0.000	10:27

08 Sim Temp = 34.0°C

Simul Ser No = MP5321
Std Sol No = 202311D
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC CHECK - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Sol.	0.298	10:33
03 Room Air	0.000	10:34
04 Std. Sol.	0.298	10:35
05 Room Air	0.000	10:35
06 Std. Sol.	0.298	10:36
07 Room Air	0.000	10:36


08 Sim Temp = 34.0°C

Simul Ser No = MP6035

Std Sol No = 202402C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:40

Test	AC	Time
01 Room Air	0.000	10:41
02 *Subject Test	INT*	10:42
03 Room Air	0.000	10:43

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 42
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:43

Test	AC	Time
01 Diagnostic	OK	10:44
02 Room Air	0.000	10:44
03 *Subject Test	RFI*	10:45
04 Room Air	0.000	10:45

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 42
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:46
02 Std. Gas	0.083	10:46
03 Room Air	0.000	10:47
04 Std. Gas	0.083	10:47
05 Room Air	0.000	10:48
06 Std. Gas	0.083	10:48
07 Room Air	0.000	10:49

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3 acn 4/29/25
cal adj needed.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
04/29/2025 10:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:49
02 Std. Gas	0.083	10:50
03 Room Air	0.000	10:50
04 Std. Gas	0.083	10:51
05 Room Air	0.000	10:51
06 Std. Gas	0.083	10:52
07 Room Air	0.000	10:52

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check - cal adj.
needed.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
06/05/2025 15:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:12
02 Std. Sol.	0.018	15:12
03 Room Air	0.000	15:13
04 Std. Sol.	0.018	15:14
05 Room Air	0.000	15:14
06 Std. Sol.	0.018	15:15
07 Room Air	0.000	15:15


08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = 202311D

County = 08

Oper No. = 133237


Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check - 0.020AC
post calibration adjustment.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
06/05/2025 15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:17
02 Std. Sol.	0.400	15:18
03 Room Air	0.000	15:18
04 Std. Sol.	0.399	15:19
05 Room Air	0.000	15:20
06 Std. Sol.	0.398	15:21
07 Room Air	0.000	15:21


08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 202312D

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC check- 0.400AC
post calibration adjustment

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
06/04/2025 12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Gas	0.080	12:42
03 Room Air	0.000	12:42
04 Std. Gas	0.080	12:42
05 Room Air	0.000	12:43
06 Std. Gas	0.080	12:43
07 Room Air	0.000	12:44

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *dry cal check #1-3*
post cal adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
06/04/2025 12:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Gas	0.080	12:45
03 Room Air	0.000	12:46
04 Std. Gas	0.080	12:46
05 Room Air	0.000	12:47
06 Std. Gas	0.080	12:47
07 Room Air	0.000	12:48

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: dry cal check #4-6
post cal. adj.

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = TOXL 8164.16.00 09/18
06/04/2025 12:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:50
02 Std. Gas	0.080	12:50
03 Room Air	0.000	12:51
04 Std. Gas	0.079	12:51
05 Room Air	0.000	12:52
06 Std. Gas	0.080	12:52
07 Room Air	0.000	12:53

Lot No = 14323080A4
Cyl No = 42
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #7-9
post cal. adj.

Form 106-I8000